



ACU/AUTY/AHRC-900(6)/ 200 /2022-23

Date: 30.12.2022

### TENDER NOTIFICATION

The Adichunchanagiri University invites **closed tenders** from eligible tenderers or bonafide licensed manufacturer or their authorised local supplier/dealer/distributor in the state of Karnataka for the procurement of Treadmill Test machine as per section I & II.

1	Name of the work	Supply of Treadmill Test (TMT) Machine to Adichunchanagiri Hospital and Research Centre, BG Nagara
2	Last date for tender submission	On or Before 13-01- 2023 up to 05:00 PM

### SECTION - I

#### Instruction to Tenderers

- The Tenderer shall send quotes in 2 bid formats (Technical and Financial bids sealed separately inside the main envelope for any or all list of items). The inner and outer sealed cover must bear the following identification
  - Tender for .....[name of service | Contract]
  - Tender Reference No.....[insert number]
  - Address to “The Registrar, Adichunchanagiri University, B.G. Nagara -571448, Nagamangala (T), Mandya (D)”
  - The tenderer who prefers to submit the tender through Post can dispatch the same through Registered Post / Speed Post or Courier so as to reach the above address on or before the due date and time specified in the Tender Notice. Tenders received after the due date and time, for what so ever reasons will not be considered and the authority, ACU BG-Nagara will not be liable or responsible for the same.
- Tender Currency:** Prices shall be quoted in Indian Rupees Only
- AMC/CMC** is subject to the Adichunchanagiri University’s norms.
- Amendment of Tender Documents:** At any time prior to the deadline for submission of tenders, the University may, for any reason, whether at its own initiative or otherwise, modify the tender documents by amendment. Adichunchanagiri University reserves all the rights to accept, reject, incorporate changes and re-tender without giving any reasons.





- **Documents Comprising the Tender:** Shall attach Brochure, Certification of the product, Bank/account details, PAN, GSTIN, Good Standing Certificate and 02 Years of ITR declaration inside the envelope and company contact details with email ID on the main envelope cover for further correspondence.
- **Tender Prices:** Prices indicated on the Price Schedule shall be entered separately I.e. the price of the goods, quoted (ex-works, ex-factory, ex-showroom, ex-warehouse, or off-the-shelf, as applicable), including all duties and sales and other taxes already paid or payable. Any Indian duties, sales and other taxes which will be payable on the goods if this Contract is awarded. Conditional tenders will not be considered.
- **Validity of the Bid:** 90 Days from the last date of submission of bid
- **Corrupt or Fraudulent practices:** The Adichunchanagiri University requires that the Tenderers, observe the highest standard of ethics during the procurement and execution of such contracts. In pursuance of this policy:
  1. will reject a proposal for award if it determines that the Tenderer recommended for award has engaged in corrupt or fraudulent practices in competing for the contract in question;
  2. will declare a firm ineligible, either indefinitely or for a stated period of time, to be awarded a university contract if it at any time determines that the firm has engaged in corrupt or fraudulent practices in competing for, or in executing, a University contract.
- **Process to be confidential:** Information relating to the examination, clarification, evaluation, and comparison of Tenders and recommendations for the award of a contract shall not be disclosed to Tenderers or any other persons not officially concerned with such process until the award to the successful Tenderer has been announced. Any effort by a Tenderer to influence the Employer's processing of Tenders or award decisions may result in the rejection of his Tender.
- **Clarification of Tenders:** To assist in the examination, evaluation, and comparison of Tenders, the Employer may, at his discretion, ask any Tenderer for clarification of his Tender, including breakdowns of unit rates. The request for clarification and the response shall be in writing or by cable, but no change in the price or substance of the Tender shall be sought, offered, or permitted except as required to confirm the correction of arithmetic errors discovered by the Employer in the evaluation of the tenders.
- **Delivery:** The successful BIDDER should commence the services as per tender document/Work or Purchase Order. For any queries/ assistance, please write to registrar@acu.edu.in or telephone to purchase section +91 -98458 35834







**SECTION -II**

**Technical Specifications**

**Treadmill Test (TMT) MACHINE WITH ECG ANALYSIS SYSTEM**

1. The Treadmill Stress Test System should be complete with acquisition of resting and stress ECG, Treadmill Unit with interface with all the protocols and provision of printing the resting as well as stress ECG and analysing the same.
2. Should acquire and analyse 12/15 simultaneous ECG Leads and display of all 12/15 leads real time Rhythm ECG on screen.
3. Should have facility of on line storage of patient ECG data. Storage patients on HDD. In addition, the storage on CD/DVD should be possible.
4. The monitor should display auto comparison of resting versus current lead of maximum ST depression separately with colour coded protocol, stage, clocks for elapsed time, total time, Target HR, Treadmill speed & grade, PVC counts/minute, warning messages & prompts, lead check torso.
5. (A) System should print comprehensive final report on a minute by minute record of ST segment changes ST segment trend plot and acceleration of ST segment.  
(B) Final request should include
  - (i) Histogram analysis of ST segment, Heart rate
  - (ii) Average median analysis of ST segment at end of each stage and at each exercise and in recovery at 1minute/3 minute/5 minute
6. System should have automatic noise free programmable treadmill
7. FDA/CE/ISI/ISO approved/certified
8. The treadmill should always start from 0 speed and has load capacity more than 200kg and speed range of 0-20 kph and elevation 0-25% and should have facility to run the self- calibration program.
9. Treadmill should have minimum 60" walking surface.
10. Shall meet IEC-60601-1-2:2001 (Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility.
11. Laser Printer to be supplied.
12. All standard accessories required for running the system to be provided.

**Dr. C.K. Subbaraya**

Registrar

Adichunchanagiri University  
B.G.Nagara-571448

