

ACU/ACN-0541/Clsd-Tender/ **2A** /2023-24Date: **18 APR 2023****TENDER NOTIFICATION**

The Adichunchanagiri University invites **closed tenders** from eligible tenderers or contractors or their authorized local supplier/dealer/distributor in the state of Karnataka for the Furnitures and Equipment for the Department of Nursing Foundation as per section I & II.

1	Name of the work	Furnitures and Equipment for the Department of Nursing Foundation, BG Nagara - 571448
2	Last date and time for receipt of tender	On or Before 02.05.2023 up to 05:00 PM

Sl. No.	Name of the work
1	Furnitures and Equipment (as per the specifications attached)

SECTION -I

SI. No	Furnitures and Equipment	Specification	Qty	Rate	Amount
1.	Fowler's cot 2 function manual	<ul style="list-style-type: none"> Should be a height adjustable, manually operated, four section cot with perforated top. The size should be approximately 2250mm length, 1015mm width and an approximate height adjustable from 520 to 820mm with $\pm 5\%$ tolerance Should have swivelling non corrosive, fibre / synthetic wheels of 125mm dia. 2 with brake and 2 without brake. Should provide with a detachable and collapsible railing made of stainless steel of minimum 19mm diameter and 18-gauge tubes housed in mild steel tubular frame, with a locking facility for raised position. Should provide with removable heavy duty saline stand made of stainless steel.304 grade. Should have a provision to fix on all four corners of the bed and also in the middle of the bed on either side Should be provided with mattress of 2 section, 4" thick PU foam of 40 density and pillows covered with soft water proof material, PVC rexin Should be supplied with mattress size should 	01		



		<p>be approximately 2010mmx910mmx100mm with 10% tolerance.</p> <ul style="list-style-type: none">• All MS components should be 7 dip tank pre-treated and epoxy powder coated with 50-60 microns.• ICU bed should be of minimum 130- 150Kg weight bearing capacity.• All MS Sheets/tubes must be CRCA• SS components wherever used should be of 304 grades.			
2.	Stretcher with I V stand	<ul style="list-style-type: none">• Size 2000 mm x 660 mm x 830 mm (L*W*H).• Frame Material Stainless Steel 304 and I V Stand.• Castor wheel Diameter 7-8 inch.• Number of Wheels 4 with diagonal locking system.• Carrying Capacity 130 - 150kg.• Contoured stretcher top should provide safety to patient handling.	02		
3	Ophthalmoscope	<ul style="list-style-type: none">• Ophthalmoscope set composed of diagnostic head mounted on a handle.• Range of lenses not smaller than -35D to +40D, adjustable in steps• Anti-reflection lens.• Ability to magnify at least five times.• Available apertures at a minimum: small, large and semi-circle, fixation star.• Colour temperature: Cool white in the range 3,500-4,000 K.• Red-free, blue, and green filters.• Handle with on/off switch.• Provided with an LED light source.• Is able to withstand frequent cleaning and disinfecting with hospital grade products.	03		




2. will declare a firm ineligible, either indefinitely or for a stated period of time, to be awarded a university contract if it at any time determines that the firm has engaged in corrupt or fraudulent practices in competing for, or in executing, a University contract.

- **Process to be confidential:** Information relating to the examination, clarification, evaluation, and comparison of Tenders and recommendations for the award of a contract shall not be disclosed to Tenderers or any other persons not officially concerned with such process until the award to the successful Tenderer has been announced. Any effort by a Tenderer to influence the Employer's processing of Tenders or award decisions may result in the rejection of his Tender.
- **Clarification of Tenders:** To assist in the examination, evaluation, and comparison of Tenders, the Employer may, at his discretion, ask any Tenderer for clarification of his Tender, including breakdowns of unit rates. The request for clarification and the response shall be in writing or by cable, but no change in the price or substance of the Tender shall be sought, offered, or permitted except as required to confirm the correction of arithmetic errors discovered by the Employer in the evaluation of the tenders.

Delivery: The successful BIDDER should commence the services as per tender document/Work or Purchase Order.

For any queries/ assistance, please write to registrar@acu.edu.in |
acu.purchase@acu.edu.in or telephone to purchase section +91 -7406907357


Head of Procurement
Adichunchanagiri University
B G Nagara -571448