



ACU/PS/AH&RC-900/7/142/2023-24

Date: 03 JAN 2024

TENDER NOTIFICATION

The Adichunchanagiri University invites **closed tenders** from eligible tenderers or bonafide licensed manufacturer or their authorised local supplier/dealer/distributor in the state of Karnataka for the procurement of Anaesthesia workstation for the Department of Renal Sciences at AHRC as per section I & II.

1	Name of the work	Supply of Anaesthesia workstation for the Department of Renal Sciences at Adichunchanagiri Hospital & Research Center, B.G. Nagara
2	Last date for tender submission	On or Before 18.01.2024 up to 05:00 PM

SECTION -I

Instruction to Tenderers

- The Tenderer shall send quotes in **2 bid formats (Technical and Financial bids sealed separately inside the main envelope for any or all list of items)** on professional business letterheads. The inner and outer sealed cover must bear the following identification
 - Tender for[name of service | Contract]
 - Tender Reference No.....[insert number]
 - Address to “The Registrar, Adichunchanagiri University, B.G. Nagara -571448, Nagamangala (T), Mandya (D)”
 - The tenderer who prefers to submit the tender through Post can dispatch the same through Registered Post / Speed Post or Courier so as to reach the above address on or before the due date and time specified in the Tender Notice. Tenders received after the due date and time, for what so ever reasons will not be considered and the authority, ACU BG-Nagara will not be liable or responsible for the same.
- Tender Currency:** Prices shall be quoted in Indian Rupees Only
- AMC/CMC (If any)** is subject to the Adichunchanagiri University’s norms.
- Warranty:** As per the Standard, preferable 03 years
- Amendment of Tender Documents:** At any time prior to the deadline for submission of tenders, the University may, for any reason, whether at its own initiative or otherwise, modify the tender documents by amendment. Adichunchanagiri University reserves all the rights to accept, reject, incorporate changes and re-tender without giving any reasons.
- Documents Comprising the Tender:** Shall attach Brochure, Certification of the product, Bank/account details, PAN, GSTIN, Good Standing Certificate and 02 Years of ITR declaration inside the envelope and company contact details with email ID on the main envelope cover for further correspondence.



7. **Tender Prices:** Prices indicated on the Price Schedule shall be entered separately I.e. the price of the goods, quoted (ex-works, ex-factory, ex-showroom, ex-warehouse, or off-the-shelf, as applicable), including all duties and sales and other taxes already paid or payable. Any Indian duties, sales and other taxes which will be payable on the goods if this Contract is awarded. Conditional tenders will not be considered.
8. **Validity of the Bid:** 90 Days from the last date of submission of bid
9. **Corrupt or Fraudulent practices:** The Adichunchanagiri University requires that the Tenderers, observe the highest standard of ethics during the procurement and execution of such contracts. In pursuance of this policy:
 1. will reject a proposal for award if it determines that the Tenderer recommended for award has engaged in corrupt or fraudulent practices in competing for the contract in question;
 2. will declare a firm ineligible, either indefinitely or for a stated period of time, to be awarded a university contract if it at any time determines that the firm has engaged in corrupt or fraudulent practices in competing for, or in executing, a University contract.
10. **Process to be confidential:** Information relating to the examination, clarification, evaluation, and comparison of Tenders and recommendations for the award of a contract shall not be disclosed to Tenderers or any other persons not officially concerned with such process until the award to the successful Tenderer has been announced. Any effort by a Tenderer to influence the Employer's processing of Tenders or award decisions may result in the rejection of his Tender.
11. **Clarification of Tenders:** To assist in the examination, evaluation, and comparison of Tenders, the Employer may, at his discretion, ask any Tenderer for clarification of his Tender, including breakdowns of unit rates. The request for clarification and the response shall be in writing or by cable, but no change in the price or substance of the Tender shall be sought, offered, or permitted except as required to confirm the correction of arithmetic errors discovered by the Employer in the evaluation of the tenders.
12. **Delivery:** The successful BIDDER should commence the services as per tender document/Work or Purchase Order. For any queries/ assistance, please write to registrar@acu.edu.in or telephone to purchase section +91 -7406907357.
13. **Penalty Clause:** Non-execution of supply order - For the reasons of failure to supply partially or completely within the stipulated time or any event of breach of contract. In case at any following stages
 1. For the delayed supply (3 days of grace period) - 5% deduction
 2. Quantity issues - 5% deduction
 3. Quality issues - 10% deduction



SECTION -II

Technical Specification of Anaesthesia Workstation for Urology department.

1. The workstation should have a built-in anaesthesia ventilator with pressure, volume controlled and paediatric modes.
2. It should be electronically controlled, pneumatically operated.
3. Should provide with adult and paediatric reusable and autoclavable lightweight tubing breathing circuit.
4. Should be able to deliver a tidal volume from 100ml to 1200ml.
5. Should have a battery backup for 30 minutes with low battery alarm and over charge protection.
6. Should have monitoring facility of airway pressure, tidal volume, frequency and oxygen concentration.
7. Should have display of at least 6 inches for set parameters and graphical display for measured parameters
8. Should have automatic self-test and leak test.
9. Anaesthesia machine should be with 3 gas supply system (O₂, N₂O, Air) with pipeline connections and reserve cylinder yokes.
10. Gas cylinder (pin indexed) yokes with sturdy clamping bars for easy handling.
11. Pin index yokes for connecting cylinders each for O₂, N₂O and air through pipeline.
12. Regulator two each for O₂ and N₂O.. N₂O should be activated only with oxygen on flow.
13. Should have pressure gauge for all gas inlets including central lines mounted on the front panel for easy visibility
14. Should have audible alarm for O₂ failure
15. N₂O supply should cut off if O₂ supply fails. (Anti-hypoxic guard).
16. Oxygen and Nitrous oxide should be linked either mechanically or pneumatically to ensure a minimum of 25% oxygen delivery at all times to avoid delivery of hypoxic mixture.
17. Should have dual cascade type flow meter for O₂ and N₂O and air calibrated in multiple scale.
18. The anaesthesia machine should have a master control ON/OFF switch.
19. Provision to mount any two selectable vaporizer with interlocking facility to allow use of only one vaporizer at a time.



20. Iso-flurane vaporizer of newer generation having specifications equivalent to tech 7/8 type to be provided.
21. Non-return cum pressure relief valve when pressure exceeds 120cm of H₂O.
22. Should have only one common gas outlet.
23. Should provide with oxygen flush switch.
24. Circle absorber with corrugated reusable breathing circuit for closed circuit system with each unit. It should be autoclavable. It should be with ventilator selector switch and circle on/off switch.
25. Should have low flow anaesthesia technique.
26. Should have a facility to connect the scavenging system.
27. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.
28. Should have a provision for mounting monitors on top of the machine and with drawers.
29. Patient Monitor Spo₂, NIBP, ECG, IBP, Temperature, ETCO₂, AGM MONITOR MAC values and NMT.
30. Should have fiber wheels and Foot brakes.
31. Standard bair circuit
32. Reservoir bag (2liters)
33. Connectors for bair circuit
34. Should be supplied with driver gas hoses with necessary attachments (colour coded).
35. Should work in 220-240Vac 50 Hz input power supply?

**Head of Procurement
Adichunchanagiri University
B G Nagara -571448**