



TENDER NOTIFICATION

The Adichunchanagiri University invites **closed tenders** from eligible tenderers or bonafide licensed manufacturer or their authorised local supplier/dealer/distributor in the state of Karnataka for the procurement of the **Rotem Machine** for Renal Science OT at AHRC as per section I & II.

1	Name of the work	Supply of Rotem Machine for Renal Science OT at Adichunchanagiri Hospital & Research Center, B.G. Nagara
2	Last date for tender submission	On or Before 23.01.2024 up to 05:00 PM

Sl. No.	Name of the consumables Particulars
1.	Rotem Machine

SECTION -I

Instruction to Tenderers

- The Tenderer shall send quotes in 2 bid formats (Technical and Financial bids sealed separately inside the main envelope for any or all list of items) on professional business letterheads. The inner and outer sealed cover must bear the following identification
 - Tender for[name of service | Contract]
 - Tender Reference No.....[insert number]
 - Address to “The Registrar, Adichunchanagiri University, B.G. Nagara -571448, Nagamangala (T), Mandya (D)”
 - The tenderer who prefers to submit the tender through Post can dispatch the same through Registered Post / Speed Post or Courier so as to reach the above address on or before the due date and time specified in the Tender Notice. Tenders received after the due date and time, for what so ever reasons will not be considered and the authority, ACU BG-Nagara will not be liable or responsible for the same.
- Tender Currency:** Prices shall be quoted in Indian Rupees Only
- AMC/CMC (If any)** is subject to the Adichunchanagiri University’s norms.
- Warranty:** As per the Standard, preferably 03 years



5. **Amendment of Tender Documents:** At any time prior to the deadline for submission of tenders, the University may, for any reason, whether at its own initiative or otherwise, modify the tender documents by amendment. Adichunchanagiri University reserves all the rights to accept, reject, incorporate changes and re-tender without giving any reasons.
6. **Documents Comprising the Tender:** Shall attach Brochure, Certification of the product, Bank/account details, PAN, GSTIN, Good Standing Certificate and 02 Years of ITR declaration inside the envelope and company contact details with email ID on the main envelope cover for further correspondence.
7. **Tender Prices:** Prices indicated on the Price Schedule shall be entered separately i.e. the price of the goods, quoted (ex-works, ex-factory, ex-showroom, ex-warehouse, or off-the-shelf, as applicable), including all duties and sales and other taxes already paid or payable. Any Indian duties, sales and other taxes which will be payable on the goods if this Contract is awarded. Conditional tenders will not be considered.
8. **Validity of the Bid:** 90 Days from the last date of submission of bid
9. **Corrupt or Fraudulent practices:** The Adichunchanagiri University requires that the Tenderers, observe the highest standard of ethics during the procurement and execution of such contracts. In pursuance of this policy:
 1. will reject a proposal for award if it determines that the Tenderer recommended for award has engaged in corrupt or fraudulent practices in competing for the contract in question;
 2. will declare a firm ineligible, either indefinitely or for a stated period of time, to be awarded a university contract if it at any time determines that the firm has engaged in corrupt or fraudulent practices in competing for, or in executing, a University contract.
10. **Process to be confidential:** Information relating to the examination, clarification, evaluation, and comparison of Tenders and recommendations for the award of a contract shall not be disclosed to Tenderers or any other persons not officially concerned with such process until the award to the successful Tenderer has been announced. Any effort by a Tenderer to influence the Employer's processing of Tenders or award decisions may result in the rejection of his Tender.
11. **Clarification of Tenders:** To assist in the examination, evaluation, and comparison of Tenders, the Employer may, at his discretion, ask any Tenderer for clarification of his Tender, including breakdowns of unit rates. The request for clarification and the response shall be in writing or by cable, but no change in the price or substance of the Tender shall be sought, offered, or permitted except as required to confirm the correction of arithmetic errors discovered by the Employer in the evaluation of the tenders.



12. **Delivery:** The successful BIDDER should commence the services as per tender document/Work or Purchase Order. For any queries/ assistance, please write to registrar@acu.edu.in or telephone to purchase section +91- 7406907357.

13. **Penalty Clause:** Non-execution of supply order - For the reasons of failure to supply partially or completely within the stipulated time or any event of breach of contract. In case at any following stages

1. For the delayed supply (3 days of grace period) - 1% deduction
2. Quantity issues - 5% deduction
3. Quality issues - 10% deduction

SECTION -II

Technical specification:-

Thrombelastograph (TEG) analyser:-

1. Complete clot analysis system with computer based feedback for individual component based clot analysis.
2. Must provide all parameters in every single reading like initiation of clotting, maximum amplitude, beginning of fibrinolysis, shear angle etc. as adynamic graph.
3. Two independent measuring channels per analyzer, up to eight channels per computer.
4. Must include all cables & parts required to connect to the computer system.
5. Cup-drive-line-synchronized, with synchronous motor.
6. Each independent measuring channel must provide complete information under one roof (not in isolation) from clot formation to clot lysis.
7. Device must perform native, citrated whole blood, heparinised & citrated heparinised samples.
8. Device should have an analog to digital converter box.
9. Software should have facility to view results of all eight channels simultaneously, allow to view /print previous data, and ease data management.
10. Monitoring software & hardware (compatible computer) must be provide in at 5 stations in the ICU & OT.
11. Device must perform the platelet mapping test & monitor the adequacy of LMWH.
12. Should perform electronic, mechanic & biological calibration.
13. Temperature control-individual temperature control for each column.
14. Measuring technique-shear elasticity of a coagulation sample, determined by motion of the pin.
15. Transducer electrical mechanical transducer of movement of torsion wire connected to the suspended pin.
16. Sample volume-360-500µl.
17. Power external power supply as per Indian standard, capable of running on battery back-up.



18. Should be ready to use at all times, initial warm-up time less than five minutes to warm sample.
19. Dimensions not more than 35cm in any dimension so that can fit a small trolley/table top.
20. Weight- about 5-7kg.
21. Adequate disposables, reagents & kits must be supplied for optimum running of the system for at least two years (Guarantee period).
22. Should be ISO/BIS approved.
23. Training must be provide to required number of hospital staff including nurses for optimum use.
24. Comprehensive maintenance warranty for at least 5years is compulsory.
25. Rate of CMC for five years after post comprehensive maintenance warranty may also quoted.

**Head of Procurement
Adichunchanagiri University
B G Nagara -571448**