



Sri Adichunchanagiri  
Shikshana Trust ®  
Mahalakshmiपुरam,  
Bengaluru-560086

SAST/CLPC/ACU-AHRC-01/01/2024-25

11-09-2024

## TENDER NOTIFICATION

The Head, CLPC, Sri Adichunchanagiri Shikshana Trust invites closed tenders from eligible tenderers or bonafide licensed manufacturer (OEM) or their authorized local supplier/ dealer/ distributor in the state of Karnataka for Procurement for Equipment's & Furniture to Adichunchanagiri Hospital and Research Center, BG nagara - 571448, Mandya as per section I & II.

01	Name of the work	Procurement of Equipment's & Furniture to Adichunchanagiri Hospital and Research Center, BG Nagara, 571448
02	Last Date for Tender Submission	On or before 18.09.2024 before 5.30 PM

### Section-1

#### Instructions to Tenderers

- 1) The Tenderer shall submit the bids (Technical & Financial bids) through the mail id: **clpchead@bgscet.ac.in** on or before the last date of tender submission (for any or all list of items) on professional business letterheads only. The details to be printed on the letter head is as follows
  - i) Tender for Procurement of Equipment's & Furniture to AHRC, BG Nagara, Mandya
  - ii) Tender Reference number.....[Insert Number]
  - iii) Address to "The HEAD, CLPC, Sri Adichunchanagiri Shikshana Trust, BGSCET Campus, Mahalakshmiपुरam, Bengaluru - 560086"
  - iv) The tenderer shall submit the original documents to this office on the last day of submission for verification who prefers to submit the tender through Post can dispatch the same through Registered post / Speed post or Couriers as to reach the above address on or before the due date and time specified in the Tender Notice. Tenders received after the due date and time, for what so ever reasons will not be considered and the authority, Head of CLPC will not be liable or responsible for the same.
- 2) **Tender Currency:** Prices shall be quoted in Indian Rupees only.
- 3) **AMC/CMC (IF ANY)** is subject to the Sri Adichunchanagiri shikshana trust's norms.
- 4) **Warranty:** As per the standards
- 5) **Amendment of tender documents:** At any time prior to the deadline of submission of tenders the trust may, for no reason, whether as its own initiative or otherwise modify the tender documents by amendment. Sri Adichunchanagiri Shikshana Trust reserves all the rights to accept, reject, incorporate changes and re-tender without giving any reasons.





Sri Adichunchanagiri  
Shikshana Trust ®  
Mahalakshmpuram,  
Bengaluru-560086

- 6) **Documents Comprising the Tender:** Shall attach Brochure, Certification of the product, Bank/account details, PAN, GSTIN, Good Standing Certificate and 02 years of ITR declaration inside the envelope and the company contact details with email id on the in the below mention format in annexure - 1.
- 7) **Tender Prices:** Prices indicated on the price schedule shall be entered separately I.e. the price of the goods, quoted (ex-works, ex-factory, ex-showroom, ex-warehouse, or off-the-shelf, as applicable), including all duties and sales and the other taxes already paid or payable. Any Indian duties, sales and other taxes which will be payable on the goods if the contract is awarded. Conditional tenders will not be considered. The bidder has to give the quotation in the below enclosed format in annexure - 2.
- 8) **Validity of the Bid:** 90 days from the last date of submission of bid.
- 9) **Corrupt or Fraudulent practices:** Sri Adichunchanagiri Shikshana Trust requires that the tenderers, observe the highest standard of ethics during the procurement and execution of such contracts. In purchase of this policy:
  - a) Will reject a proposal for award if it determines the tenderer recommended for award has engaged in corrupt or fraudulent practices in competing for the contract in question;
  - b) Will declare a firm ineligible, either indefinitely or for the stated period of time, to be awarded a university contract if it any time determines that the firm has engaged in corrupt or fraudulent practices in competing for, or in executing, a trust contract.
- 10) **Process to be confidential:** Information relating to the examination, clarification, evaluation, and comparison of tenders and recommendations for the award of contract will not be disclosed to tenderers or any other persons not officially concerned with such process until the award to the successful tenderer has been announced. Any effort by a tenderer to influence the employer's processing of tenders or award decisions may result in rejection of his tender.
- 11) **Clarification of Tenders:** To assist in the examination, evaluation, and comparison of tenders the employer may, at his discretion, ask and tenderer for clarification of his tender, including breakdowns of unit rates. The request for clarification and the response shall be writing or by cable, but no change in the price or substance of the tender shall be sought, offered, or permitted except as required to confirm the correction of arithmetic errors discovered by the employers in the evaluation of the tenders.
- 12) **Delivery:** The successful BIDDER should commence the service as per the tender document/work or purchase order. For any queries or assistance, please write to [clpthead@bgscet.ac.in](mailto:clpthead@bgscet.ac.in) or telephone to +91- 9448270489.
- 13) **Penalty Clause:** Non-execution of supply order – for the reasons of failure to supply partially or completely within the stipulated time or any event of breach of contract. In case at any following stages
  - a) For the delayed supply (3 days of grace period) – 5% deduction
  - b) Quantity issues – 5 % deduction
  - c) Quality issues – 10% deduction



Sri Adichunchanagiri  
Shikshana Trust ®  
Mahalakshmpuram,  
Bengaluru-560086

## **Section-2**

### **Technical Specification**

<b>Sl No.</b>	<b>Item Description   Particulars</b>	<b>Quantity</b>
<b>Hospital Furniture</b>		
1	Cushion Labor cot (Delivery bed)	1
<b>Instruments &amp; Equipment</b>		
1	Vacuum/ Suction machine	1
2	NST Machine with trolley	1
3	Cardiac Monitor	1





### Technical Specification of Delivery Cot

Sl. No.	Technical Specification of Delivery Cot
1	Approximate Dimensions 72"L x 30"W x24"-32"H. or equivalent.
2	Trendelenburg position adjustable.
3	Backrest Section adjustable on ratchet.
4	Leg end section can slide under the main section.
5	A pair of knee crutches which should be adjustable height
6	It should have detachable IV Stand – 2 Nos.
7	All part of the table should be made of rust-free materials.
8	It should have side railing detachable and wheel with break.
9	It should have a pair of rubber padded stainless steel lithotomy rods with ankle straps and stainless steel Bowl under " V" cut.
<b>II. Standards, Safety and Training:</b>	
1	Should be FDA or CE or BIS approved product, etc .
2	Manufacturer/Supplier should have ISO certification for quality standards.
<b>III. Documentation</b>	
1	User Manual and Service manual in English must be provided.
<b>IV Installation, Commissioning and Testing..</b>	
1	The equipment and all accessories should be transported, installed, tested and Commissioned at the Department of OBG, AHRC, B G NAGARA.
<b>V Warranty and After Sales Service:</b>	
1	The Equipment including all accessories including bought out items should be under WARRANTY for a period of THREE YEARS after successful commissioning.



## **Specification Of NST Machine with Trolley**

1. a 17" wide TFT Color LCD Screen for Fetal Heart Rate, Uterine Contractions & Fetal Movement.
2. LCD Screen should be tiltable, rotatable and can be fixed at any angle from 0-90 degree
3. It should have display modes of Trend mode, Number mode & Graph Mode.
4. It should have alarm functions in all movement.
5. It should have facilities to increase/ decrease/mute alarm sounds.
6. It should have Automatic track & hold facility.
7. It should have high sensitivity less than 1 MHz
8. FHR upper limit should be 130-240 BPM & lower limit should be 30-150 BPM.
9. It should have Fetal Heart Rate & Uterine Pressure recording system.
10. It should have automatic & Manual Fetal movement detection.
11. It must have a fetal waker device to stimulate the sleeping fetus for more accurate fetal heart rate.
12. It should have inbuilt thermal recorder.
13. It should have battery backup.
14. It should have a built-in memory for 450 hrs. to save patient graph data.
15. It should have Twin Fetal Monitoring system.
16. FHR & UC (Uterine Contraction) Probes should have a slot guided connecting mechanism to the main unit.
17. It should have print speed of 1, 2, 3cm/min.
18. It should have automatic paper feeding system.
19. It should have safety standard certificate like CE or FDA.

It should be supplied with standard accessories

- US probe/FHR probe - 1 nos.
- Toco Probe - 1 no.
- Event Maker Jack - 1 no.
- Print paper - 5 rolls
- Power adapter & Cord - 1 no.
- Ultrasound Gel - 4 no.
- Probe Belt - 2sets

20. Trolley to keep the fetal monitor/ machine - 1 no.





### Technical Specification of patient monitor

1. Should have TFT/LCD/Touchscreen display with at least 10 - 15 inches and at least 4-5 wave forms touch screen and numeric display simultaneously
2. The waveform should be user selectable with preset user defined settings for varying care locations.
3. Monitor should have in built Lithium-ion type battery for 2 Hrs continuous operation in case of mains failure
4. Should have keys for quick access to main functions.
5. Should be able to monitor Neonatal ECG, SPO2, NIBP and Respiration Rate.
6. 5 Lead ECG monitoring.
7. Respiration & Apnea alarm
8. Manual, Auto and STAT mode for NIBP monitoring and ranges should be 30 to 230mmHg.
9. Pulse Oximeter (SPO2) with Plethysmograph & Pulse strength indicator With Variable pitch with change in SpO2. Should have low perfusion compensation and motion tolerance technology.
10. Should have separate volume control for beep sound for QRS and alarm sound.
11. Should provide following accessories
  - a. Reusable adult 3/5 lead ECG cable set (Clip type) – 3 nos.
  - b. Reusable adult and Neonatal SPO2 finger probes 1 each, Extension cable - 1 each
  - c. NIBP cuffs for standard Child- 1 each, Neonatal-1 each.
12. Should work on 200-240V AC/ 50Hz with inbuilt rechargeable battery.
13. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission/ FDA
14. Facility for last 24 hours full disclosure with wave forms trend review & facility for patient data entry
  - a. Monitor stand as per the following specification should be available
  - b. Monitor stand - extruded Aluminum, powder coated.
  - c. Load bearing capacity 20 kgs approximately.



Sri Adichunchanagiri  
Shikshana Trust ®  
Mahalakshmpuram,  
Bengaluru-560086

**Annexure - 1**  
**PARTICULARS OF THE BIDDER**

<b>Sr. No</b>	<b>Description</b>	<b>Details (to be filled by the responder to the Bid)</b>
1	Name of the company	
2	Official address	
3	Phone No. And Fax No.	
4	Corporate Headquarters Address	
5	Phone No. And Fax No.	
6	Web Site Address	
7	Details of Company's Registration (Please enclose copy of the company registration document)	
8	Name of Registration Authority	
9	Registration Number and Year of Registration	
10	ISO certifications and its validity	
11	GST registration No.	
12	Permanent Account Number (PAN)	
13	Company's Revenue for last 3 years (Year wise)	
14	Company's net worth for the last year	
15	Bank Details (Name, Account no., Branch, IFSC, MICR)	





Sri Adichunchanagiri  
Shikshana Trust @  
Mahalakshmiपुरam,  
Bengaluru-560086

## Annexure – 2


The Bidder has to quote the rate in the Item Data available online with this bid. Details to be filled up for price bid are as below:

**The price shall be inclusive of all taxes (inclusive of GST) under the relevant Laws of India.**

SL. No	Particular	Amount In Rs. ( Inclusive of All the taxes)
1	Total Cost for Procurement & supply of Equipment's & Furniture to AHRC	
Total in Rs and in words –		

Cost related to Supply and Installation as per Items mentioned in the Compliance sheet for technical proposal.

SL. No	Item Description	Quantity	Unit Rate		Total Commercial Cost in Rs.
			Base price in Rs.	% GST/HSN	
1.	Cushion Labour/Delivery cot	1			
2.	Vacuum/ Suction machine	1			
3.	NST Machine with trolley	1			
4.	Cardiac Monitor	1			
Total in Rs					

  
**HEAD, CLPC**  
**SRI ADICHUNCHANAGIRI**  
**SHIKSHANA TRUST**  
**BANGALORE – 560086**