

SAST/CLPC/ACU- SACCP/47/2024-25

02-01-2025

## TENDER NOTIFICATION

The Head, CLPC, Sri Adichunchanagiri Shikshana Trust invites closed tenders from eligible tenderers or bonafide licensed manufacturer (OEM) or their authorized local supplier/ dealer/ distributor in the state of Karnataka for the Procurement of Lab Equipments to SAC College of Pharmacy (SACCP), ACU, BG Nagara- 571448 as per section I & II.

01	Name of the work	Procurement of Lab Equipments to SAC College of Pharmacy (SACCP), ACU, BG Nagara- 571448
02	Last Date for Tender Submission	On or before 10.1.2025 before 5.30 PM

## Section-1


## Instructions to Tenderers

- 1) The Tenderer shall submit the bids (Technical & Financial bids) through the mail id: clpchead@bgscet.ac.in on or before the last date of tender submission (for any or all list of items) on professional business letterheads only. The details to be printed on the letter head is as follows
  - i) Tender for Procurement of Lab Equipments to SAC College of Pharmacy (SACCP), ACU, BG Nagara- 571448
  - ii) Tender Reference number......[Insert Number]
  - iii) Address to ""The HEAD, CLPC, Sri Adichunchanagiri Shikshana Trust, BGSCET Campus, Mahalakshmipuram, Bengaluru – 560086"
  - iv) The tenderer shall submit the original documents to this office on the last day of submission for verification who prefers to submit the tender through Post can dispatch the same through Registered post / Speed post or Couriers as to reach the above address on or before the due date and time specified in the Tender Notice. Tenders received after the due date and time, for what so ever reasons will not be considered and the authority, Head of CLPC will not be liable or responsible for the same.
- 2) Tender Currency: Prices shall be quoted in Indian Rupees only.
- 3) AMC/CMC (IF ANY) is subject to the Sri Adichunchanagiri shikshana trust's norms.
- 4) Warranty: 3Years
- 5) Amendment of tender documents: At any time prior to the deadline of submission of tenders the trust may, for no reason, whether as its own initiative or otherwise modify the tender documents by amendment. Sri Adichunchanagiri Shikshana Trust reserves all the rights to accept, reject, incorporate changes and re-tender without giving any reasons.
- 6) Documents Comprising the Tender: Shall attach Brochure, Certification of the product, Bank/account details, PAN, GSTIN, Good Standing Certificate and 02 years of ITR declaration inside the envelope and the company contact details with email id on the in the below mention format in annexure - 1.

1 | Page

B.G. Nagara - 571448, NH-75, Nagamangala Tq., Mandya Dist., Karnataka, India, Tel: 08234 287285 | Email: Info@acu.edu.in, registrar@acu.edu.in | www.acu.edu.in

KSHANA TRUST



- 7) **Tender Prices:** Prices indicated on the price schedule shall be entered separately I.e. the price of the goods, quoted (ex-works, ex-factory, ex-showroom, ex-warehouse, or off-the-shelf, as applicable), including all duties and sales and the other taxes already paid or payable. Any Indian duties, sales and other taxes which will be payable on the goods if the contract is awarded. Conditional tenders will not be considered. The bidder has to give the quotation in the below enclosed format in annexure 2.
- 8) Validity of the Bid: 90 days from the last date of submission of bid.
- 9) Corrupt or Fraudulent practices: Sri Adichunchanagiri Shikshana Trust requires that the tenderers, observe the highest standard of ethics during the procurement and execution of such contracts. In purchase of this policy:
  - a) Will reject a proposal for award if it determines the tenderer recommended for award has engaged in corrupt or fraudulent practices in competing for the contract in question;
  - b) Will declare a firm ineligible, either indefinitely or for the stated period of time, to be awarded a university contract if it any time determines that the firm has engaged in corrupt or fraudulent practices in competing for, or in executing, a trust contract.
- 10) **Process to be confidential:** Information relating to the examination, clarification, evaluation, and comparison of tenders and recommendations for the award of contract will not be disclosed to tenderers or any other persons not officially concerned with such process until the award to the successful tenderer has been announced. Any effort by a tenderer to influence the employer's processing of tenders or award decisions may result in rejection of his tender.
- 11) Clarification of Tenders: To assist in the examination, evaluation, and comparison of tenders the employer may, at his discretion, ask and tenderer for clarification of his tender, including breakdowns of unit rates. The request for clarification and the response shall be writing or by cable, but no change in the price or substance of the tender shall be sought, offered, or permitted except as required to confirm the correction of arithmetic errors discovered by the employers in the evaluation of the tenders.
- 12) **Delivery:** The successful BIDDER should commence the service as per the tender document/work or purchase order. For any queries or assistance, please write to <a href="mailto:clpchead@bgscet.ac.in">clpchead@bgscet.ac.in</a> or telephone to +91- 9448270489.
- 13) **Penalty Clause:** Non-execution of supply order for the reasons of failure to supply partially or completely within the stipulated time or any event of breach of contract. In case at any following stages
  - a) For the delayed supply (3 days of grace period) 5% deduction
  - b) Quantity issues 5 % deduction
  - c) Quality issues 10% deduction

HEAD, CLICA




# Section-2

### **Technical Specification**

Sl No	Material/Service (Lab Equipments)	Technical Evaluation - Details of requirements	QTY
1 Deep Freezer		To Store Samples and tissue that has to be stored at low temperature Upto -80°C, 200 to 250 litre capacity	1
High precision Analytical weighing balance  To weigh very small quantity of samples Weighing range 220g, readability 0.0001g		To weigh very small quantity of samples Weighing range 220g, readability 0.0001g	1
3	Centrituge upto 15000 rpm, -20 C to 40 C		1
Refrigerating Centrifuge  To Centrifuge blood samples or of upto 15000 rpm, -20°C to 40°C  Specifications for Freeze Dryer -60 to -80 Freeze Dryer Standard chamber with 8 port manifold Stoppering Final condenser tempn-60 Vacuum Degree (Pa) <10 Condenser volume (L) 6.5 Freeze drying area (m2) 0.09 Ice condenser capacity (Kg/24h Qty of shelf 3 or more Material loading capacity/shelf Material loading capacity (ml) 9 Freeze drying time (h) 24 System should have USB Interf Drying Chamber(standard) Tra Vacuum Pump(standard) Stand Speed: 2L/S Control System Microprocessor, Power supply (V/Hz) 220V/50H Exterior dimension (WxDxH mm Should have provided CE certific Drying chamber (standard) Vacu system Standard model: Speed: 2L/S Microprocessor, tou Ultimate vacuum (Pa) Noise (dB) requirement Electrical voltage Ele Power (Kw) Size information Econ		-60 to -80 Freeze Dryer Stoppering Chamber Standard chamber with 8 port manifold Stoppering chamber Final condenser tempn-60 Vacuum Degree (Pa) <10 Condenser volume (L) 6.5 Freeze drying area (m2) 0.09 Ice condenser capacity (Kg/24h) 3 Qty of shelf 3 or more Material loading capacity/shelf (ml) 300 Material loading capacity (ml) 900 Freeze drying time (h) 24 System should have USB Interface Drying Chamber(standard) Transparent acrylic Vacuum Pump(standard) Standard model: 2XZ-2, Speed: 2L/S Control System Microprocessor, touch screen Power supply (V/Hz) 220V/50Hz,60Hz Exterior dimension (WxDxH mm) 600*540*820/880 Should have provided CE certification Drying chamber (standard) Vacuum pump Control	1

**3** | Page


B.G. Nagara - 571448, NH-75, Nagamangala Tq., Mandya Dist., Karnataka, India.
Tel: 08234 287285 | Email: info@acu.edu.in, registrar@acu.edu.in | www.acu.edu.in
SHIKSHANA TRUST
BANGALORE - 560000



Vendor should have provided user list and installation certificates   Vendor should have Local offices In Bangalore for service & Spares Support	MINIC	7116. 46 111/14/16	B.G. Nagara - 571448	
1   1   1   1   1   1   1   1   1   1			certificates Vendor should have Local offices In Bangalore for	
Immunoshop Lab Basic Cat no RT-6900  1    Stability chamber   HUMIDITY CHAMBER/STABILITY CHAMBER   Inner SS 304 & Outer MS   Temperature Range: 10°C above ambient to 60°C   1    But	5	7.7	UV-Vis spectrophotometer	1
Tissue homogenizer  Stability chamber  Inner SS 304 & Outer MS Temperature Range: 10°C above ambient to 60°C  1  PH METER MICROPROCESSOR BASED 3 POINT CALIBRATION WITH AUTO ATC Model: LMPH-10 HSN CODE: 90278090  1  Centrifuge  Cilnical Centrifuge Model: C-854/6 Brand: REMI  1  Digital Viscometer  DIGITAL ROTATIONAL VISCOMETER MODEL 60 Model: LMDV60 HSN CODE: 90278010  1  Tissue homogenizer  Model: RQT-127 A/D Brand: REMI  1  Mechanical stirrer  Catalogue number RQ 100 Plus Laboratory Stirrer Make: REMI  1  LIV sephinet  THO selection with an appearance of the component of the c	6	_	Immunoshop Lab Basic	1
B Digital Ph meter 3 POINT CALIBRATION WITH AUTO ATC Model: LMPH-10 HSN CODE: 90278090 1  Centrifuge Centrifuge Model: C-854/6 Brand: REMI 1  Digital Viscometer DIGITAL ROTATIONAL VISCOMETER MODEL 60 Model: LMDV60 HSN CODE: 90278010 1  Tissue homogenizer Model: RQT-127 A/D Brand: REMI 1  Mechanical stirrer Catalogue number RQ 100 Plus Laboratory Stirrer Make: REMI 1  LIV ochinet Tional Action Live Live Live Live Live Live Live Live	7	Stability chamber	Inner SS 304 & Outer MS	1
9 Centrifuge Model: C-854/6 Brand: REMI 1  10 Digital Viscometer DIGITAL ROTATIONAL VISCOMETER MODEL 60 Model: LMDV60 HSN CODE: 90278010 1  11 Tissue homogenizer Model: RQT-127 A/D Brand: REMI 1  12 Mechanical stirrer Catalogue number RQ 100 Plus Laboratory Stirrer Make: REMI 1	8	Digital Ph meter	3 POINT CALIBRATION WITH AUTO ATC Model: LMPH-10	1
Digital Viscometer Model: LMDV60 HSN CODE: 90278010  Tissue Homogeniser Model: RQT-127 A/D Brand: REMI  Mechanical stirrer Catalogue number RQ 100 Plus Laboratory Stirrer Make: REMI	9	Centrifuge	Model: C-854/6	1
11 Tissue homogenizer Model: RQT-127 A/D Brand: REMI  12 Mechanical stirrer Catalogue number RQ 100 Plus Laboratory Stirrer Make: REMI	10	Digital Viscometer	Model: LMDV60	1
12 Mechanical stirrer Stirrer Make: REMI	11	Tissue homogenizer	Model: RQT-127 A/D	1
13 UV cabinet TLC plate viewer- UV cabinet 1	12	Mechanical stirrer	Stirrer	1
	13	UV cabinet	TLC plate viewer- UV cabinet	1

B.G. Nagara - 571448, NH-75, Nagamangala Tq., Mandya Dist., Karnataka, India. Tel: 08234 287285 | Email: Info@acu.edu.in, registrar@acu.edu.in | www.acu.edu.in

> SHIKSHANA TRUST BANGALORE - 56008






14	Drying oven	HOT AIR OVEN – MEMMERT TYPE WITH AIR CIRCULATION FAN Inner SS and Outer MS - Temperature Range: 50°C to 250°C Make: Anamatrix HSN CODE: 8417	1
15	Rotary evaporator	CAT no. 0010003502 (IKA- Germany) RV 3 ECO	1
16	Magnetic stirrer with heating and ceramic heating plate  Magnetic stirrer with heating and ceramic heating plate  CERAMIC HOTPLATE MAGNETIC STIRRER Model: LMMS-5LC		1
17	Homogenizer	Tissue Homogeniser Model: RQT-127 A/D Brand: REMI	1
18	Incubator Shaker	ROTARY INCUBATOR Cat no: RS-18 Plus Make: REMI	1
19	Vacuum Pump	CAT no.: N96 DIAPHRAGM VACUUM PUMP 7LPM,130mbar	1
20	Analytical balance	ANALYTICAL BALANCE 220g/0.1mg Model: ATX224 (ATX-R) Brand: Shimadzu HSN Code: 90160010	1
21	Sonicator	ULTRASONIC BATH 10L Brand :Labman HSN Code : 84562000	1
22	Shimadzu UV-VIS Spectrophotometer	High visibility color touch panel with stylus, Best in class scan speed up to 29,000 nm/min for high speed kinetic studies, Wide wavelength range of 1,100 nm to 190 nm, High resolution 1 nm spectral bandwidth over entire wavelength range, Wavelength accuracy of +0.5nm for D2 spectral line, Wavelength reproducibility of + 0.1nm, Wide Photometric range of -4 to +4 Abs and 0 to 400 %T, High baseline flatness.	1

B.G. Nagara - 571448, NH-75, Nagamangala Tq., Mandya Dist., Karnataka, India. Tel: 08234 287285 | Email: info@acu.edu.in, registrar@acu.edu.ln:

SHIKSHANA TRUST BANGALORE - 560086






Exitation / Emission: Concave, blazed holographic with lower stray light, 1300 grooves / mn, Wavelength Range: 200 to 900nm and 0 order Spectrofluorophoto Spectral Bandwidth: Excitation: 1.5, 3, 5, 10, 15, 20nm				
Emission: 1, 3, 5, 10, 15, 20nm Wavelength Accuracy: Within ±1.0nm, Wide varity of Spectral Techniques, Measurement, Quantitation, Photometric, Time Course, Data	23	Spectrofluorophoto meter	blazed holographic with lower stray light, 1300 grooves / mn, Wavelength Range: 200 to 900nm and 0 order Spectral Bandwidth: Excitation: 1.5, 3, 5, 10, 15, 20nm Emission: 1, 3, 5, 10, 15, 20nm Wavelength Accuracy: Within ±1.0nm, Wide varity of Spectral Techniques, Measurement,	1

HEAD, CLPC SRI ADICHUNCHANAGIR SHIKSHANA TRUST BANGALORE - 560086



# Annexure - 1 PARTICULARS OF THE BIDDER

Sr. No		Details (to be filled by the responder to the Bid)
1	Name of the company	
2	Official address	
3	Phone No. And Fax No.	
4	Corporate Headquarters Address	Helmanders (Klain Missian Since) And Manual Commence of the Co
5	Phone No. And Fax No.	
6	Web Site Address	
7	Details of Company's Registration (Please enclose copy of the company registration document)	
8	Name of Registration Authority	
9	Registration Number and Year of Registration	
10	ISO certifications and its validity	
11	GST registration No.	
12	Permanent Account Number (PAN)	
13	Company's Revenue for last 3 years (Year wise)	
14	Company's net worth for the last year	
15	Bank Details (Name, Account no., Branch, IFSC, MICR)	





#### Annexure - 2

The Bidder has to quote the rate in the Item Data available online with this bid. Details to be filled up for price bid are as below:

The price shall be inclusive of all taxes (inclusive of GST) under the relevant Laws of India.

SL. No	Particular	Amount In Rs.  (Inclusive of All the taxes)
1	Total Cost for the Procurement of Lab Equipments to SAC College of Pharmacy (SACCP), ACU, BG Nagara- 571448	
Total in	Rs and in words –	

RI ADICHUNG AN AOA SHIKSHAM RUST BANGALORE - 560086