



SAST/CLPC/ACU-AHRC/63 /2024-25

23-01-2025

TENDER NOTIFICATION

The Head, CLPC, Sri Adichunchanagiri Shikshana Trust invites closed tenders from eligible tenderers or bonafide licensed manufacturer (OEM) or their authorized local supplier/ dealer/ distributor in the state of Karnataka for the **Procurement of OT Table for Urology and Neuro Surgery at BGS MCH Hospital, Nagarur** as per section I & II.

01	Name of the work	Procurement of OT Table for Urology and Neuro Surgery at BGS MCH Hospital, Nagarur
02	Last Date for Tender Submission	On or before 27.01.2025 before 5.30 PM

Section-1

Instructions to Tenderers

- 1) The Tenderer shall submit the bids (Technical & Financial bids) through the mail id: **clpthead@bgscet.ac.in** on or before the last date of tender submission (for any or all list of items) on professional business letterheads only. The details to be printed on the letter head is as follows
 - i) Tender for Procurement of OT Table for Urology and Neuro Surgery at BGS MCH Hospital, Nagarur.
 - ii) Tender Reference number.....[Insert Number]
 - iii) Address to "The HEAD, CLPC, Sri Adichunchanagiri Shikshana Trust, BGSCET Campus, Mahalakshmpuram, Bengaluru - 560086"
 - iv) The tenderer shall submit the original documents to this office on the last day of submission for verification who prefers to submit the tender through Post can dispatch the same through Registered post / Speed post or Couriers as to reach the above address on or before the due date and time specified in the Tender Notice. Tenders received after the due date and time, for what so ever reasons will not be considered and the authority, Head of CLPC will not be liable or responsible for the same.
- 2) **Tender Currency:** Prices shall be quoted in Indian Rupees only.
- 3) **AMC/CMC (IF ANY)** is subject to the Sri Adichunchanagiri shikshana trust's norms.
- 4) **Warranty:** 3 Years
- 5) **Amendment of tender documents:** At any time prior to the deadline of submission of tenders the trust may, for no reason, whether as its own initiative or otherwise modify the tender documents by amendment. Sri Adichunchanagiri Shikshana Trust reserves all the rights to accept, reject, incorporate changes and re-tender without giving any reasons.
- 6) **Documents Comprising the Tender:** Shall attach Brochure, Certification of the product, Bank/account details, PAN, GSTIN, Good Standing Certificate and 02 years of ITR declaration inside the envelope and the company contact details with email id on the in the below mention format in annexure - 1.



- 7) **Tender Prices:** Prices indicated on the price schedule shall be entered separately i.e. the price of the goods, quoted (ex-works, ex-factory, ex-showroom, ex-warehouse, or off-the-shelf, as applicable), including all duties and sales and the other taxes already paid or payable. Any Indian duties, sales and other taxes which will be payable on the goods if the contract is awarded. Conditional tenders will not be considered. The bidder has to give the quotation in the below enclosed format in annexure - 2.
- 8) **Validity of the Bid:** 90 days from the last date of submission of bid.
- 9) **Corrupt or Fraudulent practices:** Sri Adichunchanagiri Shikshana Trust requires that the tenderers, observe the highest standard of ethics during the procurement and execution of such contracts. In purchase of this policy:
 - a) Will reject a proposal for award if it determines the tenderer recommended for award has engaged in corrupt or fraudulent practices in competing for the contract in question;
 - b) Will declare a firm ineligible, either indefinitely or for the stated period of time, to be awarded a university contract if it any time determines that the firm has engaged in corrupt or fraudulent practices in competing for, or in executing, a trust contract.
- 10) **Process to be confidential:** Information relating to the examination, clarification, evaluation, and comparison of tenders and recommendations for the award of contract will not be disclosed to tenderers or any other persons not officially concerned with such process until the award to the successful tenderer has been announced. Any effort by a tenderer to influence the employer's processing of tenders or award decisions may result in rejection of his tender.
- 11) **Clarification of Tenders:** To assist in the examination, evaluation, and comparison of tenders the employer may, at his discretion, ask and tenderer for clarification of his tender, including breakdowns of unit rates. The request for clarification and the response shall be writing or by cable, but no change in the price or substance of the tender shall be sought, offered, or permitted except as required to confirm the correction of arithmetic errors discovered by the employers in the evaluation of the tenders.
- 12) **Delivery:** The successful BIDDER should commence the service as per the tender document/work or purchase order. For any queries or assistance, please write to clpchead@bgscet.ac.in or telephone to +91- 9448270489.
- 13) **Penalty Clause:** Non-execution of supply order - for the reasons of failure to supply partially or completely within the stipulated time or any event of breach of contract. In case at any following stages
 - a) For the delayed supply (3 days of grace period) - 5% deduction
 - b) Quantity issues - 5 % deduction
 - c) Quality issues - 10% deduction



Section-2

Technical Specification

SL. No	Particulars	Total Quantity in Nos.	UOM
OT TABLE FOR UROLOGY			
1.	OT Table as per technical specification	1	Nos
2.	Padded arm rest with straps - pair with damps	1	set
3.	Anaesthesia screen with clamps	1	set
4.	Side supports: pair with clamps	1	set
5.	Shoulder supports: pair with clamps	1	set
6.	Knee crutches: pair with damps	1	set
7.	X-ray cassette tray with pushing rod	1	Nos
8.	Gel Flat button chest roll (medium)	2	Nos
9.	Gel prone positioner (feet)	1	Nos
10.	Gel prone head rest (large)	1	Nos
11.	X-Ray Top with 9" Urology extension	1	Nos
12.	Metal Drain pan	1	Nos
13.	Power lift stirrup set with side rail clamp	1	set
14.	Anti-Trendelenburg restraint strap	3	Nos
15.	Light weight transfer board	1	Nos
16.	Detachable Head Rest, tilting 15 deg Up and 90 deg Down, manually operated.	1	Nos
17.	Detachable pair of lithotomy stirrups with padded knee crutches mounted on ball socket joints and stainless-steel lithotomy stirrups mounted on universal rotary clamps made of stainless steel.	1	Nos
18.	Detachable pair of split leg boards, swiveling out 0 degree to 90 degrees.	1	Nos



OT TABLE FOR UROLOGY	
Sl. No	Technical Specification
1	Electromechanical Operating Tables are tables for performing surgical procedures and it works with electrical power.
2	Operational Requirements
a	OT Table is required for Urology surgeries of all types. and should have X-Ray translucent tops.
3	Technical Specifications
a	Four / Five section table top with divided foot section (split leg section)
b	Table should be independently maneuverable, can be moved with or without patient and stopped in any position.
c	Modular OR tabletop, consisting of 4/5 sections, with mounting points on the end sides for the adaptation of head rest, leg board & various attachments.
d	Table top should permit x-ray penetration and fluoroscopy with full length X-ray cassette tunnel accessible from either end.
e	The main central section padding should be removable easily without the use of tools.
f	The tabletop should be a special composite material to minimize radiation hazard to Surgeons and OR staff. The table top should be free of any screws or nut bolts, and should be crevices free, for better infection control.
g	All table positioning, i.e., height, back section adjustment, lateral tilt, Cranial- caudal traversing or longitudinal movement, Trendelenburg and anti-Trendelenburg, should be operated through handset and standby control
h	Column cover should be stainless steel. The operating tabletop is equipped with side rails of Stainless steel enabling accessory adaption. The base exterior of the table should be molded plastic, for easy maintenance and rust-free long life.
i	Mattress should be radiolucent and of PU make and suitable for fluoroscopy. Preferably Mattress for each section of the table should be separate
j	Zero level button
k	Table should have following movements:-
l	Measurements: (+/-5% tolerance allowed in the sizes/measurements/angles/dimensions)
m	Height (without pad): 600mm – 1000mm Height:



n	Trendelenburg / reverse Trendelenburg 30° / 30°
o	Lateral tilt left / right 20° / 20°
p	Back section up 70° / down 40°
q	Longitudinal shift: 300 mm or more
r	Foot section adjustment: - 90 to 0 degree, detachable
s	All motorized movements must be precise, smooth and jerk free.
4	Dimensions :
a	Table length should be 1900 mm minimum
b	Table Width should be 550 mm minimum
c	Maximum patient weight: 250 kgs in all positions
d	Table should have wired remote control handset for all movements of the table. There should be an auxiliary emergency operating panel on the table base for independent operation of the movements of the table. Table should also have a wireless remote control operating from outside the sterile zone.
5	Standards & Safety
	Should be BIS certified for the quoted model OR
a	should meet IEC 60601-1, IEC 60601-1-2 & IEC 60601-2-46 standards and should submit valid test report for the quoted model from NABL accredited laboratory/ from the labs in their country of origin in case of foreign manufacturer
b	It Should have import/manufacturing license from Central licensing Authority or State licensing authority of CDSCO for the quoted model and copy of valid license should be submitted in line with the prevailing rules/guidelines/circular from CDSCO
6	System Configuration Accessories, spares and consumables
a	All the accessories should be from the same manufacturer (except gel pad & light weight transfer board) and same quality standards as asked for table. The price of the following accessories should be quoted separately and will be taken for price bid evaluation and ranking purpose.
b	Padded arm rest with straps - pair with damps - 1 set
c	Anaesthesia screen with clamps - 1 set
d	Side supports: pair with clamps - 1 set
e	Shoulder supports: pair with clamps - 1 set
f	Knee crutches: pair with damps - 1 set
g	X-ray cassette tray with pushing rod - 1 set
h	The machines should be supplied with the following accessories for operating prone position
l	X-Ray Top with 9" Urology extension - 1 no
m	Metal Drain pan - 1 no



n	Pneumatic lift stirrup set with side rail clamp - 1 set
o	Anti-Trendelenburg restraint strap - 3 no
p	Detachable Head Rest, tilting 15 deg Up and 90 deg Down, manually operated. - 1 no
q	Detachable pair of lithotomy stirrups with padded knee crutches mounted on ball socket joints and stainless steel lithotomy stirrups mounted on universal rotary clamps made of stainless steel. - 1 no
r	Detachable pair of split leg boards, swivelling out 0 degree to 90 degrees. - 1 no
s	Light weight transfer board - 1 no
7	Power Supply
a	Power input: 220-240V/ 50 Hz AC single phase fitted with appropriate Indian plugs and sockets.
b	Inbuilt battery backup for 5 hrs or equivalent
8	Training
a	Comprehensive training for staff of user department and support services till familiarity with the system.
9	Demonstration is must. Demonstration shall be arranged at HITES office, Noida or in Delhi NCR region & bidder should provide demonstration accordingly.

SL. No	Particulars	Total Quantity in Nos.	UOM
OT Table for NEURO SURGERY			
1.	OT Table as specified	1	Nos
2.	Arm board - 2	2	Nos
3.	Body strap- 3	3	Nos
4.	Anesthesia screen with clamps- 2	2	Nos
5.	Side supports with clamps - 2	2	Nos
6.	Clamp, rotary- 4 pc	4	Nos
7.	Clamp, circular - 4 pc	4	Nos
8.	Accessories stand, mobile on castors- 1 pc	1	Nos
9.	Arm support, perplex -2 pc	2	Nos
10.	Clamp for locking X Ray cassette -1	1	Nos
11.	Accessories for operating in prone - a radiolucent	1	Set



	Wilson frame, gel pad for head, chest & feet - 1 each		
12.	3 pin skull clamps with Adult & Pediatric reusable pins (2 sets each) and horse shoes for adult & pediatric sizes with table attachment (1 each)	1	Set
13.	Cross bar with clamps for sitting position - 1 set	1	set

Sl. No	Technical Specification
1	Multipurpose Electro Mechanical with manual/electric override mobile Table with divided leg section suitable for all Neuro surgery of all types.
A	General operating table features:
1	Full-length radio-translucent top.
2	Table should be independently maneuverable, can be moved with or without patient and stopped in any position. Table should have motorized floor locking by remote control.
3	4 or 5 sections tabletop, which should be made of a special scratch resistant, hardwearing and easy to clean material. Base column cover to be made of 100% stainless steel alloy and stainless steel.
4	Horizontal alignment of all tilts of the tabletop via the zero-position function on the hand control
5	Mattress should be of high quality that spans tabletop break for improved patient support. Its depth should be 50mm. Mattress must be Latex free.
6	The main central section padding should be removable easily without the use of tools.
7	Table should not have a thread/sharp edge for ensuring proper cleaning and user safety.
8	Column cover should be stainless steel. The operating tabletop is equipped with side rails enabling accessory adaption. The base exterior of the table should be molded plastic, for easy maintenance and rust-free long life. There should be no exposed mild steel parts, so that table is rust free throughout its lifetime.
9	The robust handset should offer controls for Trend. /Reverse Trend, Lateral Tilt, Flexion/ Extension, horizontal sliding and Height functions. One button auto levelling should be available
10	Brakes, 4nos Wheels
11	Table should have a narrow base allowing optimum access of c arm and greater stability
12	It should have a stable construction with lockable castors via a central foot pedal/ Hand control



B Technical specifications:	
1	Following movements of the table must be motorized, controlled from the remote handset :
2	Height (without pad): 600 mm - 1000mm
3	Trendelenburg / reverse Trendelenburg 25° / 25°
4	Lateral tilt left / right 15° / 15°
5	Back section up 70° / down 40°
6	Automatic 0 position function of Table top from all tilts of Trendelenburg and lateral.
7	Longitudinal shift: 200 mm or more
8	Floor locking and unlocking of table base.
9	All motorized movements must be precise, smooth and jerk free.
C Dimensions:	
1	Table length should be 1900mm minimum
2	Width of the tabletop should be 500mm ($\pm 10\%$) without side rails
3	Table should have wired remote control handset for all movements of the table. There should be an auxiliary emergency operating panel on the table base for independent operation of the movements of the table. Table should also have a wireless remote control, so that table can be operated from outside the sterile zone.
4	Table should have an inbuilt Battery backup with visual charge level indicator showing percentage of charge left in the battery.
5	Permissible load of the operating table: Maximum patient weight: 250 kgs
6	Should be BIS certified for the quoted model OR should meet IEC 60601-1, IEC 60601-1-2 & IEC 60601-2-46 standards and should submit valid test report for the quoted model from NABL accredited laboratory/ from the labs in their country of origin in case of foreign manufacturer

[Handwritten signature]


Annexure - 1
PARTICULARS OF THE BIDDER

Sr. No	Description	Details (to be filled by the responder to the Bid)
1	Name of the company	
2	Official address	
3	Phone No. And Fax No.	
4	Corporate Headquarters Address	
5	Phone No. And Fax No.	
6	Web Site Address	
7	Details of Company's Registration (Please enclose copy of the company registration document)	
8	Name of Registration Authority	
9	Registration Number and Year of Registration	
10	ISO certifications and its validity	
11	GST registration No.	
12	Permanent Account Number (PAN)	
13	Company's Revenue for last 3 years (Year wise)	
14	Company's net worth for the last year	
15	Bank Details (Name, Account no., Branch, IFSC, MICR)	



Annexure - 2

The Bidder has to quote the rate in the Item Data available online with this bid. Details to be filled up for price bid are as below:

The price shall be inclusive of all taxes (inclusive of GST) under the relevant Laws of India.

SL. No	Particular	Amount In Rs. (Inclusive of All the taxes)
1	Total Cost for the Procurement of OT Table for Urology and Neuro Surgery at BGS MCH Hospital, Nagarur	
Total in Rs and in words -		

Cost related to Supply and Installation as per Items mentioned in the Compliance sheet for technical proposal.

SL. No	Particulars	Total Quantity In nos.	Unit Rate in Rs	Total Cost in Rs.
OT TABLE FOR UROLOGY BOQ				
1	OT Table as per technical specification	1		
2	Padded arm rest with straps - pair with damps	1		
3	Anaesthesia screen with clamps	1		
4	Side supports: pair with clamps	1		
5	Shoulder supports: pair with clamps	1		
6	Knee crutches: pair with damps	1		
7	X-ray cassette tray with pushing rod	1		
8	Gel Flat button chest roll (medium)	2		
9	Gel prone positioner (feet)	1		
10	Gel prone head rest (large)	1		
11	X-Ray Top with 9" Urology extension	1		



12	Metal Drain pan	1		
13	Power lift stirrup set with side rail clamp	1		
14	Anti-Trendelenburg restraint strap	3		
15	Light weight transfer board	1		
16	Detachable Head Rest, tilting 15 deg Up and 90 deg Down, manually operated.	1		
17	Detachable pair of lithotomy stirrups with padded knee crutches mounted on ball socket joints and stainless-steel lithotomy stirrups mounted on universal rotary clamps made of stainless steel.	1		
18	Detachable pair of split leg boards, swiveling out 0 degree to 90 degrees.	1		
OT Table for NEURO SURGERY BOQ				
1	OT Table as specified	1		
2	Arm board - 2	2		
3	Body strap- 3	3		
4	Anesthesia screen with clamps- 2	2		
5	Side supports with clamps - 2	2		
6	Clamp, rotary- 4 pc	4		
7	Clamp, circular - 4 pc	4		
8	Accessories stand, mobile on castors- 1 pc	1		
9	Arm support, perplex -2 pc	2		
10	Clamp for locking X Ray cassette -1	1		
11	Accessories for operating in prone - a radiolucent Wilson frame, gel pad for head, chest & feet - 1 each	1		
12	3 pin skull clamps with Adult & Pediatric reusable pins (2 sets each) and horse shoes for adult & pediatric sizes with table attachment (1 each)	1		
13	Cross bar with clamps for sitting position - 1 set	1		
Total in Rs				
GST 18% in Rs.				
Grand Total Amount in Rs.				