



### TENDER NOTIFICATION

The Head, CLPC, Sri Adichunchanagiri Shikshana Trust invites closed tenders from eligible tenderers or bonafide licensed manufacturer (OEM) or their authorized local supplier/ dealer/ distributor in the state of Karnataka for the **Procurement of Equipment's and instruments for the Department of Plastic Surgery, Reconstructive and Aesthetic Surgery at Adichunchanagiri Hospital and Research Center, BG Nagara, Mandya (District) - 571448.** as per section I & II.

01	Name of the work	Procurement of Equipment's and instruments for the Department of Plastic Surgery, Reconstructive and Aesthetic Surgery at Adichunchanagiri Hospital and Research Center, BG Nagara, Mandya (District) - 571448.
02	Last Date for Tender Submission	On or before 14.02.2025 before 5.30 PM

### Section-1

#### Instructions to Tenderers

- 1) The Tenderer shall submit the bids (Technical & Financial bids) through the mail id: **clpchead@bgscet.ac.in** on or before the last date of tender submission (for any or all list of items) on professional business letterheads only. The details to be printed on the letter head is as follows
  - i) Tender for Procurement of Equipment's and instruments for the Department of Plastic Surgery, Reconstructive and Aesthetic Surgery at Adichunchanagiri Hospital and Research Center, BG Nagara, Mandya (District) - 571448.
  - ii) Tender Reference number.....[Insert Number]
  - iii) Address to "The HEAD, CLPC, Sri Adichunchanagiri Shikshana Trust, BGSCET Campus, Mahalakshampuram, Bengaluru - 560086"
  - iv) The tenderer shall submit the original documents to this office on the last day of submission for verification who prefers to submit the tender through Post can dispatch the same through Registered post / Speed post or Couriers as to reach the above address on or before the due date and time specified in the Tender Notice. Tenders received after the due date and time, for what so ever reasons will not be considered and the authority, Head of CLPC will not be liable or responsible for the same.
- 2) **Tender Currency:** Prices shall be quoted in Indian Rupees only.
- 3) **AMC/CMC (IF ANY)** is subject to the Sri Adichunchanagiri shikshana trust's norms.
- 4) **Warranty:** As per the standards **13 years.**





- 5) **Amendment of tender documents:** At any time prior to the deadline of submission of tenders the trust may, for no reason, whether as its own initiative or otherwise modify the tender documents by amendment. Sri Adichunchanagiri Shikshana Trust reserves all the rights to accept, reject, incorporate changes and re-tender without giving any reasons.
- 6) **Documents Comprising the Tender:** Shall attach Brochure, Certification of the product, Bank/account details, PAN, GSTIN, Good Standing Certificate and 02 years of ITR declaration inside the envelope and the company contact details with email id on the in the below mention format in annexure - 1.
- 7) **Tender Prices:** Prices indicated on the price schedule shall be entered separately I.e. the price of the goods, quoted (ex-works, ex-factory, ex-showroom, ex-warehouse, or off-the-shelf, as applicable), including all duties and sales and the other taxes already paid or payable. Any Indian duties, sales and other taxes which will be payable on the goods if the contract is awarded. Conditional tenders will not be considered. The bidder has to give the quotation in the below enclosed format in annexure - 2.
- 8) **Validity of the Bid:** 90 days from the last date of submission of bid.
- 9) **Corrupt or Fraudulent practices:** Sri Adichunchanagiri Shikshana Trust requires that the tenderers, observe the highest standard of ethics during the procurement and execution of such contracts. In purchase of this policy:
  - a) Will reject a proposal for award if it determines the tenderer recommended for award has engaged in corrupt or fraudulent practices in competing for the contract in question;
  - b) Will declare a firm ineligible, either indefinitely or for the stated period of time, to be awarded a university contract if it any time determines that the firm has engaged in corrupt or fraudulent practices in competing for, or in executing, a trust contract.
- 10) **Process to be confidential:** Information relating to the examination, clarification, evaluation, and comparison of tenders and recommendations for the award of contract will not be disclosed to tenderers or any other persons not officially concerned with such process until the award to the successful tenderer has been announced. Any effort by a tenderer to influence the employer's processing of tenders or award decisions may result in rejection of his tender.
- 11) **Clarification of Tenders:** To assist in the examination, evaluation, and comparison of tenders the employer may, at his discretion, ask and tenderer for clarification of his tender, including breakdowns of unit rates. The request for clarification and the response shall be writing or by cable, but no change in the price or substance of the tender shall be sought, offered, or permitted except as required to confirm the correction of arithmetic errors discovered by the employers in the evaluation of the tenders.
- 12) **Delivery:** The successful BIDDER should commence the service as per the tender document/work or purchase order. For any queries or assistance, please write to [clpchead@bgscet.ac.in](mailto:clpchead@bgscet.ac.in) or telephone to +91- 8123707324.
- 13) **Penalty Clause:** Non-execution of supply order - for the reasons of failure to supply partially or completely within the stipulated time or any event of breach of contract. In case at any following stages
  - a) For the delayed supply (3 days of grace period) - 5% deduction
  - b) Quantity issues - 5 % deduction
  - c) Quality issues - 10% deduction



**Section-2****Technical Specification**

**Equipment's and Instruments for the Department of Plastic Surgery, Reconstructive and Aesthetic Surgery List:**

Sl. No	Description	Qty	Amount
<b>I. Plastic Surgery Major OT</b>			
1	Ultrasound Bovie	1	
2	Versajet Debridement System III	1	
3	Never Stimulator	1	
4	Tendon Grafting and Tendon Transfer Set	1	
5	Limb Washing basin trolley	1	

Sl. No	Description	Qty	Amount
<b>II. Laser Equipment</b>			
1	Vertical Diode 1600W triple wavelength	1	
2	Active Q switch Vertical	1	
3	C20 Laser vertical 60W	1	
4	HydraFacial Super Vertical	1	
5	Micro needling RF Machine vertical	2	
6	RF cautery 4.2MHz	1	
7	Centrifuge with brushless rotor	1	
8	Derma chair 3 motor for 2 Nos	1	

Sl. No	Description	Qty	Amount
<b>III. National Medical Council Requirement</b>			
1	Ear lobe Clamp for 2 Nos	2	





2	Facelift Set	1	
3	NPWT Machine	1	
4	Orthognathic, Head and neck set	1	
5	Mirrorless Digital Camera	1	
6	Tripod for camera	1	
7	Flash lighting system with stands	1	
<b>TOTAL AMOUNT</b>			

### Plastic Surgery Major OT

#### Technical Specifications:

#### 1. Ultrasound Bovie: Ultrasonic Shears & Scalpel Qty -1

- Energy generator console/device
- Universal connector
- High resolution display
- Software updates
- Advanced Bipolar
- Advanced Ultrasound
- Class 1 device
- 100 -240V, 50/60Hz. 500VA

(a)Hand Piece

Ultrasonic Shears, Scalpel set

Curved jaw/tip

Used for  $\leq$  5mm vessels & Lymphatics

(b) Foot switch

Multifunction

Waterproof Certification

#### 2. Hydro-Surgery Dissection System / Versajet Debridement System III – Qty -1

(a) Console





Simple User interface

Adequate displays

Power cord

100-250V

IP certification (water)

(b) Hand Piece

Different Curved angles

Easy connections to console

Multiple water jet channels at the tip

(c) Foot Switch

Multi-functional

IP certification (water)

### 3. Nerve Stimulator Set

#### (a) Stimulator for direct and US guided (deep) stimulation Qty -1

Battery

Long Cable for device connection

Set of Needles for deep stimulation

User manual

0.00 to 5.00 mA

0.00 to 1.00 mA

Frequency adjustment:

1Hz/2Hz etc

Stimulus duration adjustment:

0.05ms – 0.10ms etc upto 1.00 ms

#### (b) Optional Accessory: Cutaneous electrical stimulator (Pen) Qty-1





4. Tendon Grafting and Tendon Transfer Set:

1	Self-retaining retractor	1
2	Bunnell Tendon passer	1
3	Tendon stripper	1
4	Steinmann Tendon holding forceps	2
5	Brand Tendon tunnelling forceps	1
6	Tendon interlacing forceps	1
7	Carroll tendon tunnelling forceps	1
8	Metzenbaum scissors 15cm TC Straight	2
9	Iris scissors 12cm straight	2
10	Iris curved scissors 12cm	2
11	Scalpel handle No 3	1
12	Mayo Hegar Needle holder	2
13	Mosquito forceps curved 12cm	8
14	Allis tissue forceps	2

5. Limb Washing:

(a) Stainless steel tray Qty -2

3ft x 1&1/2 ft

3 inches depth

(b) Stainless steel tray Qty -2

2ft x 1&1/2 ft

3 inches depth

(c) Stainless steel wire mesh: Qty -2





3ft x 1& ½ ft

1.5mm

Material grade 304

Smooth surface

**(d) Stainless steel wire mesh: Qty -2**

2ft x 1& ½ ft

1.5mm

Material grade 304

Smooth surface

**(e) Strong Plastic stool to support weight of the leg 1square feet ½ feet height Qty -4**

**(f) Stainless steel basins Qty- 4**

Round

1&1/2 feet diameter

20 gauge or lesser

**LASER EQUIPMENT'S**

**A) Vertical Diode Laser Qty -1**

- Triple wavelength Diode laser hair removal Advanced system
- Power 1600 Watt,
- Probe size 12\*20
- Cooling System
- Nose Tip
- Vertical Model





**B) Active Q switch Vertical Qty -1**

- Wavelength: 1064nm 532nm Standard: 585nm,650nm Optional
- Energy: 600mj (1064nm): 300mj (532nm)
- Peak Power: 1064nm 1.2GW: 532nm 0.6GW
- Frequency: 1-10Hz
- Zoom Spot Size: 2-10mm Adjustable
- Pulse Width: 450ps
- Beam Profile: Top Hat Beam
- Light Guiding System: 7 joints Arm
- Aiming Beam: Diode 655 nm (Red),

**C) CO2 Laser Qty -1**

- The type of laser: 10600nm Carbon dioxide laser
- Focus spot diameter:  $\leq 0.1$ mm
- Maximum Output Power: 70w
- Pulse width: 0.1~10ms (0.1ms step)
- Repeat time:1~20 times (1 step)
- Scan Shape: Square; Rectangle; Circle; Triangle; Hexagon; Ellipse; Line
- Maximum scan area: 20×20mm
- Scan Patterns: Standard; Random; Scatter
- Screen: Large multifunction screen, preferably touch-screen
- Voltage: AC 220V 50Hz

**D) HydraFacial Qty -1**

- Functions: skin rejuvenation, wrinkle removal, facial cleaning, facial deep peeling, skin tightening, lifting, acne treatment
- Ultrasonic frequency: 1 MHz-3Mhz





- RF type: Bipolar and Quadrupole
- Quantity of Hydro dermabrasion facial bottle: Type A B C D
- Hydro head facial: 6 hard heads and 2 polished heads
- Oxygen jet, facial deep cleaning, skin activating
- Display: Multi-function display screen, preferably touchscreen
- The temperature scope of the Cold & hot hammer: 0-45°C
- The power of the machine: <350W
- The hydro dermabrasion handle pressure:90Kpa
- The hydro dermabrasion handle flow:30L

**E) Micro Needling RF Machine Qty -2**

- Dimension(cm): 40\*38\*80
- Weight (net weight):20 kg
- RF Frequency: 5Mhz
- RF Output: 80w
- Work mode: MFR, SFR
- MFR Needle depth: 0.3-3 mm adjustable (0.1mm step)
- MFR Delay Time: 1-10 ms
- SFR Pulse Time: 1000ms-3000ms adjustable (100ms step)
- RF Energy intensity: 1-10 level
- Invasive head area(cm<sup>3</sup>): 11,1.5 15,2-2
- Led:650nm/5m WX1
- Non-invasive head: 36 pin/2\*2 cm<sup>2</sup>
- Input Voltage: 110/220V, 50/60Hz





**F) RF Cautery Qty -1**

- Digital Display
- Power Input: 230V /50 Mhz
- Frequency: 2Mhz
- Power Output:
- Cut: 120 Watts on 400 ohm
- Coagulation: 100 watts on 400 ohm
- Bipolar: 80 watts on 100 ohm
- Portable

**G) Remi Centrifuge with Brushless Rotator Qty -1**

- Max. Speed: 6000 RPM
- Max. RCF: 5070 g
- Max. Capacity: 400ml
- Digital timer range: 0-99
- Low Noise (preferable  $\leq 60$  or 70 dB)
- Supply: 220-240Volts 50 Hz.

**H) Derma chair Qty -2**

- Bed Size :80-95 Inch Length
- Height Position: Adjustable (average 32 Inch)
- Lowest Position:  $\leq 24$  Inch
- Adjustable Angle of head position: Straight Detachable
- Adjustable Angle of Back rest part: 0 Degree – 80 Degree (Up)
- Adjustable Angle of Leg rest part:180 Degree- 75 Degree (Down)
- Adjustable Angle of Seat Setting: Straight  $< 24$  Inch to  $> 32$  Inch Up
- Net weight: 60 - 100 Kg (Chair Weight)





- Maximum loading weight: 200 Kilograms or more
- Height of Head part: Detachable & Adjustable
- Adjustable Length of Head part: extendable minimum 11 Inch
- Adjustable Length of Leg rest part: 22 Inch or more

**I) Microdermabrasion Vertical Qty -1**

- Rated Voltage: AC220V±20%, 50Hz AC110V±20%, 60Hz
- Handpieces: Crystal Dermabrasion Handle, Diamond Dermabrasion Handle, Jet Spray Handle
- Rated Power: 110W
- Microcurrent Power: 10W
- Vacuum Pressure: 0-75Kpa
- Portable
- 

**J) Mesotherapy Machine Qty -1**

- Voltage - 220v / 50 Hz
- Display – Large multifunction screen display, preferably touchscreen
- Ultrasound frequency - 1 Mhz, 3Mhz
- Treatment Time - 20 minutes
- End Of Therapy – Buzzer/Alarm
- Duty Cycle - 5~100%
- 

**K) Coolsculpting Crayo + EMS**

Sl no	Particulars	Specification
1	Handle configuration	4 handles
2	Screen Size	Large with Touchscreen
3	Electromagnetic energy	0-5T
4	Stimulus pulse shape	Biphasic wave
5	Pulse	300 (microseconds)
6	Cooling system	Water cycle semiconductor refrigeration





**Specifications For Facelift and Blepharoplasty Instruments**

SL NO	DESCRIPTION	Length in mm	Quantity
<b>1</b>	<b>SKIN HOOKS SINGLE</b>		
(i)	Round Skin Hook- Fine	150	1
(ii)	Round Skin Hook- sharp	150	1
(iii)	Round Skin Hook Medium- sharp	190	1
(iv)	Round Skin Hook Light Weight- sharp	150	1
(v)	Flat Skin Hook – Short – sharp	125	1
(vi)	Flat Skin Hook – Long – sharp	150 / 170	1
<b>2</b>	<b>SKIN HOOKS DOUBLE</b>		
(i)	Double Skin hook - 4 / 5 Mm – sharp		1
(ii)	Double Skin hook - 6 Mm – sharp		1
(iii)	Double Skin hook - 10 / 12 Mm- sharp		1
(iv)	Double Skin hook -blunt	160	1
<b>3</b>	<b>ENDOSCOPIC ELEVATORS</b>		
(i)	Endoscopic Elevators Small -head size 4.5 mm	250	1
(ii)	Endoscopic Elevators Medium Head size 6mm	250	1
(iii)	Endoscopic Elevators Large	250	1





	head size 8mm		
(iv)	Endoscopic Elevators Malleable - Malleable head size 10mm	310	1
(v)	Endoscopic Light system and accessories		1
<b>4</b>	<b>KILLIANS SEPTAL ELEVATOR</b>		
(i)	Killians Septal Elevator- Straight Both ends straight	250	1
(ii)	Killians Septal Elevator- Curved Both ends curved	250	1
(iii)	Killians Septal Elevator- With Suction One end straight with suction	250	1
<b>5</b>	<b>JOSEPH SKIN ELEVATOR</b>		
(i)	Joseph Skin Elevator Long Blade size-30x5mm	165	1
(ii)	Joseph Skin Elevator Short Blade size 20 x 5mm	165	1
6	Freer elevator	180	1
7	Pierce elevator	215	1
8	Cottle elevator	190	1
9	Masing graduated elevator	185	1
10	Farabeuf elevator straight	185	1
11	Farabeuf elevator curved	185	1
	<b>RETRACTORS</b>		
12	Kilner Ala Retractor 9 mm- head size 6mm	95.25	1
13	Double ended Kilner ala retractor 8mm /10mm Head size 8mm and 10mm	114.3	1
14	Face Lift Retractor	175	1





	4 prong retractor 5mm Gap between hooks		
15	Coronal Brow Lift 7 Prongs Retractor, 18 Cm 7 prong retractors 2 outer and 5 inner prongs with gap of 10mm 72 mm head	175	1
16	Desmarres Lid Retractor Small 7mm Head size 7 mm	140	1
17	Desmarres Lid Retractor Small 11mm Head size 11mm	140	1
18	Desmarres Lid Retractor Small 13mm Head size 13 mm	140	1
19	Aufricht retractor wide Blade size 12 x 45mm	185	1
20	Aufricht retractor narrow Blade size 7 x 50mm	185	1
21	Aufricht walled retractor Blade size 7 x 45mm One end retractor and other end double hook	140	1
22	Senn Double End Retractor -Claw Type One end sharp claw and other end flat	150	1
23	Senn Miller Retractor One end sharp claw and other end curved	150	1
24	Senn Double End Retractor -plain Both ends plain	150	1
25	Senn Double End Retractor -fine mini Both ends fine and plain	147.5	1
26	Desmarres Lid Retractor Medium 11 mm Head size 11mm	140	1
			1
27	FIBER OPTIC Light System console		1
28	Aufricht Retractor With Fiberoptic Retractor with fiber optic connection		1
29	Killian nasal speculum with Fiberoptic		1





	Retractor with fiber optic connection		
	<b>MEASURING INSTRUMENTS</b>		
30	RULER (Metal) 6inch/15cm		2
31	CALLIPER 20MM 20 mm measuring scale with marking on both sides		2
	Suction Tips		
32	Suction Tips 7" Diameter - 2.5 mm	125	2
33	Suction Tips 9" Diameter - 3 mm	125	2
34	Suction Tips 12" Diameter - 4 mm	125	2
	Scalpel Handles & Probes		
35	Knife Handle No 3 Al Handle Anadoised handle for holding scalpel blades no: 11 AND 15		2
36	Knife Handle No 4		2
37	Bowmans Double Ended Probes 1&2, 3&4, 5&6, 7&8: Set of 4		1





	<b>FORCEPS</b>		
38	Adsons Forceps Toothed 5" Toothed	127	1
39	Adsons Forceps Non-Toothed 5" Plain	127	1
40	Adsons Forceps Toothed 6" Toothed	150	1
41	Debakey Forceps 6" Atraumatic toothed	160	1
42	Debakey Forceps 7" Atraumatic toothed	180	1
43	Debakey Forceps 8 Atraumatic toothed "	210	1
44	Lid forceps HEAD SIZE 15 MM WITH FINE SERRATIONS	140	1
45	Iris Forceps Straight Straight plain	120	1
45	Iris Forceps Curved Curved	120	1
46	Jewellers Forceps Curved with Lock Curved with lock	105	1
47	St.Martins Forceps 3.5" Straight non toothed	85	1
48	Chalazion Forceps HEAD SIZE 15 MM WITH FINE SERRATIONS	105	1





49	Chalazion Forceps Fenestrated 24 mm Oval head 24mm fenestrated	95	1
50	Mosquito Forceps 5 " Straight Straight	135	1
51	Mosquito Forceps 5 " Curved Curved	135	1
52	Adsons Forceps Cross Serrated Cross serrated head	127	1
53	Cartilage Forceps 5" Serrated and grooved	127	1
54	Bayonet Forceps 6" OR Bayonet 5 ½" Bayonet shaped	150	1
55	Tilleys Dressing Forceps		1
56	Lucs Forceps small	127	1
57	Lucs Forceps Medium	165	1
58	Lucs Forceps Large	190	1
59	Asch Forceps	241.3	1
60	Walsham Forceps Left	215	1
61	Walsham Forceps Right	215	1
62	Baby Walsham Forceps	215	1
63	Nasal Septal Forceps Straight (OR) Nasal Septal Forceps Angled		1
64	Baby Mosquito Forceps Straight	101	1
65	Baby Mosquito Forceps Curved	101	1





SCISSORS			
66	Iris Scissors 4 1/2" – Straight Straight	104	1
67	Iris Scissors 4 1/2" – Curved Curved	104	1
68	Iris Scissors Curved 4" Curved	101	1
69	Iris Super Cut Scissors STR / CD Serrated	115	1
70	Metzenbaum Scissors 7" Curved curved	180	1
	Metzenbaum Scissors 6" Curved	150	1
71	Tenotomy Scissors – Straight	127	1
72	Tenotomy Scissors – Curved	127	1
73	Tenotomy Scissors Curved Mini Slightly curved	101	1
74	Face Lift Scissors – Straight Long Blade  BLADE SIZE 50 MM SERRATED STRAIGHT	172	1
75	Face Lift Scissors – Curved Long Blade  BLADE SIZE 50 MM SERRATED CURVED	172	1
76	Strabismus Scissors straight	110	1
77	Strabismus scissors curved	110	1
78	Trespat Scissors Oval shaped blade	185	1





79	TC Iris Scissors curved TC tip	116	1
80	Kilner Scissor Straight	127	1
81	Kilner Scissor Curved	127	1
82	Fomon Angular Scissors Angled	165	1
83	Fomon Angular Scissors Curved	165	1
84	Cartilage Scissors Straight Head 35mm	190	1
85	Cartilage Scissors Curved Head 35mm	190	1
86	Turbinectomy Scissors Angled	165	1
87	Turbinectomy Scissors Curved	165	1
88	Converse Angled Scissor 4" Angled	101	1
89	Aufricht Scissors/ plain or serrated		1
90	Columella Scissors 5" Tip curved	127	1
91	Columella Scissors 6" Tip curved	150	1
92	Columella Scissors 7" Tip curved	175	1
93	Tc Tip Tenotomy Scissors Curved	127	1
94	Tc Tip Metz 6" Scissors Curved	150	1
95	Gold cut Scissors	200	1
96	Stevens Super Cut Scissors Straight	127	1





	Serrated		
97	Stevens Super Cut Scissors Curved Serrated	127	1
	Needle Holder		
98	Needle Holder 5" Grooved jaw	130	1
99	5" Rhyder Needle Holder Fine jaw	127	1
100	Webster Needle Holder Plain jaw	127	1
101	Gillies Needle Holder Needle holder with scissors	127	1
102	Tc Tip Needle Holder 5" With TC tip	127	1
	Others		
103	Ptosis Clamp with lock 25 mm head size	110	1
104	Ptosis Knife 13 mm blade slightly curved	194	1
105	Rhytidoplasty Marker One end plain with reverse hooks and other end curved with three hooks	150	1
<b>Orthognathic, Head &amp; Neck Set</b>			
	Hooks		





	Serrated		
97	Stevens Super Cut Scissors Curved Serrated	127	1
	Needle Holder		
98	Needle Holder 5" Grooved jaw	130	1
99	5" Rhyder Needle Holder Fine jaw	127	1
100	Webster Needle Holder Plain jaw	127	1
101	Gillies Needle Holder Needle holder with scissors	127	1
102	Tc Tip Needle Holder 5" With TC tip	127	1
	Others		
103	Ptosis Clamp with lock 25 mm head size	110	1
104	Ptosis Knife 13 mm blade slightly curved	194	1
105	Rhytidoplasty Marker One end plain with reverse hooks and other end curved with three hooks	150	1
<b>Orthognathic, Head &amp; Neck Set</b>			
	Hooks		





106	Double Hook Sharp Medium 10/12 mm		4
	Elevator		
107	Bristow elevator straight		1
108	Bristow elevator curved		1
109	Periosteal elevator with fiber handle		Set of 3
110	Zygoma Elevator (Rowes)		1
111	Warwick James elevator Right		1
112	Warwick James elevator Left		1
	Retractor		
113	Rayne Mandibular Retractor		1
114	Chin Retractor		1
115	Forked Ramus Retractor (Ramus with bifurcated		1
116	Channel Retractor		1
117	Kilner Retractor "c" shape double ended small		1
118	Kilner Retractor "c" shape double ended medium		1
119	Kilner Retractor "c" shape double ended Large		1
120	Austin Tissue Retractor Regular		1
121	Langenbeck Retractor small		1





122	Langenbeck Retractor Medium		1
123	Langenbeck Retractor Large		1
124	Condyle Retractor Double hook type		1
125	Self-Retaining Mollison Retractor ( Medium)		1
126	Rowes Orbital Floor Retractor right & left set		1
127	Superior Ramus Retractor Smith		1
128	Soft Tissue / Reverse Langenbeck Retractor		1
	<b>FORCEPS</b>		
129	Rowes Maxilla Disimpaction Forceps: Left		1
130	Rowes Maxilla Disimpaction Forceps: Right		1
131	Hayton William Forceps Forward Traction Forceps		1
132	Reduction Bone Holding Forceps Pointed		1
133	TMJ Spreading Forceps Double ended		1
134	Aluminium box 12.5" x 8.5" x 3"		1
135	Fully equipped Cranio-maxillary fixation set for reconstruction Includes Plates, Magnetic Screws, drillbits, drivers, plate bender & cutter, guides, storage box		1
136	20-gauge Dental wire		2
137	22-gauge dental wire		2





138	24-gauge dental		2
139	Arch bar		2
140	<b>Mirrorless Digital Camera</b>  Automatic  High build quality  Inbuilt flash  Full Frame  E-mount  Back illuminated  4k, 1080, etc  Wide range of ISO sensitivity		1
141	Tripod for camera		1
142	<b>Lens system</b>  Compatible with the Camera device  28-60mm zoom		1
143	Lighting System Complete with umbrella & stand		1
144	Electrical extension board		2
145	<b>Camera Memory card reader with Multiple size slots</b>		1





	Micro USB, USB 2.0, USB 3.0 & higher C type reversible port SD, SDHC, SDXC, Micro SD, fast transfer Multiple sizes memory card compatible slots		
146	<b>Portable Negative Pressure Wound Therapy (NPWT)</b>		
	Device Class	C	
	Operating Modes	Continuous, Intermittent	
	Pressure settings	75 mmHg, -100 mmHg, -125mmHg	01
	Power Source	AC-DC Adapter-Input AC 100-240V, 50-60 Hz 5V DC	
	Battery	4.2V Dc,5200mAh	
147	<b>Ear lobe Clamps - Multi- size clamps</b>		02



**Annexure - 1****PARTICULARS OF THE BIDDER**

<b>Sr. No</b>	<b>Description</b>	<b>Details (to be filled by the responder to the Bid)</b>
1	Name of the company	
2	Official address	
3	Phone No. And Fax No.	
4	Corporate Headquarters Address	
5	Phone No. And Fax No.	
6	Web Site Address	
7	Details of Company's Registration (Please enclose copy of the company registration document)	
8	Name of Registration Authority	
9	Registration Number and Year of Registration	
10	ISO certifications and its validity	
11	GST registration No.	
12	Permanent Account Number (PAN)	
13	Company's Revenue for last 3 years (Year wise)	
14	Company's net worth for the last year	
15	Bank Details (Name, Account no., Branch, IFSC, MICR)	



**Annexure - 2**

The Bidder has to quote the rate in the Item Data available online with this bid. Details to be filled up for price bid are as below:

**The price shall be inclusive of all taxes (inclusive of GST) under the relevant Laws of India.**

SL. No	Particular	Amount In Rs. (Inclusive of All the taxes)
1	Total Cost for the Procurement of Equipment's and instruments for the Department of Plastic Surgery, Reconstructive and Aesthetic Surgery at Adichunchanagiri Hospital and Research Center, BG Nagara, Mandya (District) - 571448.	
Total in Rs and in words -		

*[Handwritten Signature]*  
4/2/25

HEAD, CLPC  
ADICHUNCHANAGIRI  
SHIKSHANA TRUST  
BANGALORE - 560086