



SAST/CLPC/ACU-AHRC/155 /2025-26

24-11-2025

**TENDER NOTIFICATION**

The Head, CLPC, Sri Adichunchanagiri Shikshana Trust invites closed tenders from eligible tenderers or bonafide licensed manufacturer (OEM) or their authorized local supplier/ dealer/ distributor in the state of Karnataka for the **Procurement Equipment's for the department of Medical Gastroenterology at Adichunchanagiri Hospital and Research Center, BG Nagara - 571448, Mandya (District)** as per section I & II.

01	Name of the work	Procurement Equipment's for the department of Gastroenterology at Adichunchanagiri Hospital and Research Center, BG Nagara - 571448, Mandya (District)
02	Last Date for Tender Submission	On or before 04.12.2025 before 5.30 PM

**Section-1**

**Instructions to Tenderers**

- 1) The Tenderer shall submit the bids (Technical & Financial bids) through the mail id: **clphead@bgscet.ac.in** on or before the last date of tender submission (for any or all list of items) on professional business letterheads only. The details to be printed on the letter head is as follows
  - i) Tender for Procurement Equipment's for the department of Gastroenterology at Adichunchanagiri Hospital and Research Center, BG Nagara - 571448, Mandya (District)
  - ii) Tender Reference number.....[Insert Number]
  - iii) Address to "The HEAD, CLPC, Sri Adichunchanagiri Shikshana Trust, BGSCET Campus, Mahalakshmpuram, Bengaluru - 560086"
  - iv) The tenderer shall submit the original documents to this office on the last day of submission for verification who prefers to submit the tender through Post can dispatch the same through Registered post / Speed post or Couriers as to reach the above address on or before the due date and time specified in the Tender Notice. Tenders received after the due date and time, for what so ever reasons will not be considered and the authority, Head of CLPC will not be liable or responsible for the same.
- 2) **Tender Currency:** Prices shall be quoted in Indian Rupees only.
- 3) **AMC/CMC (IF ANY)** is subject to the Sri Adichunchanagiri shikshana trust's norms.
- 4) **Warranty:** 3 Years.
- 5) **Amendment of tender documents:** At any time prior to the deadline of submission of tenders the trust may, for no reason, whether as its own initiative or otherwise modify the tender documents by amendment. Sri Adichunchanagiri Shikshana Trust reserves all the rights to accept, reject, incorporate changes and re-tender without giving any reasons.

- 6) **Documents Comprising the Tender:** Shall attach Brochure, Certification of the product, Bank/account details, PAN, GSTIN, Good Standing Certificate and 02 years of ITR declaration inside the envelope and the company contact details with email id on the in the below mention format in annexure - 1.
- 7) **Tender Prices:** Prices indicated on the price schedule shall be entered separately I.e. the price of the goods, quoted (ex-works, ex-factory, ex-showroom, ex-warehouse, or off-the-shelf, as applicable), including all duties and sales and the other taxes already paid or payable. Any Indian duties, sales and other taxes which will be payable on the goods if the contract is awarded. Conditional tenders will not be considered. The bidder has to give the quotation in the below enclosed format in annexure – 2.
- 8) **Validity of the Bid:** 90 days from the last date of submission of bid.
- 9) **Corrupt or Fraudulent practices:** Sri Adichunchanagiri Shikshana Trust requires that the tenderers, observe the highest standard of ethics during the procurement and execution of such contracts. In purchase of this policy:
  - a) Will reject a proposal for award if it determines the tenderer recommended for award has engaged in corrupt or fraudulent practices in competing for the contract in question;
  - b) Will declare a firm ineligible, either indefinitely or for the stated period of time, to be awarded a university contract if it any time determines that the firm has engaged in corrupt or fraudulent practices in competing for, or in executing, a trust contract.
- 10) **Process to be confidential:** Information relating to the examination, clarification, evaluation, and comparison of tenders and recommendations for the award of contract will not be disclosed to tenderers or any other persons not officially concerned with such process until the award to the successful tenderer has been announced. Any effort by a tenderer to influence the employer's processing of tenders or award decisions may result in rejection of his tender.
- 11) **Clarification of Tenders:** To assist in the examination, evaluation, and comparison of tenders the employer may, at his discretion, ask and tenderer for clarification of his tender, including breakdowns of unit rates. The request for clarification and the response shall be writing or by cable, but no change in the price or substance of the tender shall be sought, offered, or permitted except as required to confirm the correction of arithmetic errors discovered by the employers in the evaluation of the tenders.
- 12) **Delivery:** The successful BIDDER should commence the service as per the tender document/work or purchase order. For any queries or assistance, please write to [clpthead@bgscet.ac.in](mailto:clpthead@bgscet.ac.in) or telephone to +91- 8123707324.
- 13) **Penalty Clause:** Non-execution of supply order – for the reasons of failure to supply partially or completely within the stipulated time or any event of breach of contract. In case at any following stages
  - a) For the delayed supply (3 days of grace period) – 5% deduction
  - b) Quantity issues – 5 % deduction
  - c) Quality issues – 10% deduction

**Section-2****Technical Specification**

SL. No	Particulars	Total Quantity in Nos.
1.	Anesthesia Workstation	1
2.	Cardiac Monitor	1
3.	ERCP Patient Table (OT Table)	1

**1. Anesthesia Workstation****Technical Specification of Anesthesia Workstation for the Department of Gastroenterology.**

1. The workstation should have a built-in anesthesia ventilator with pressure, volume controlled and pediatric modes.
2. It should be electronically controlled, pneumatically operated.
3. Should provide with adult and pediatric reusable and autoclavable light weight tubing breathing circuit.
4. Should be able to deliver a tidal volume from 100ml to 1200ml.
5. Should have a battery back up for 30 minutes with low battery alarm and over charge protection.
6. Should have monitoring facility of airway pressure, tidal volume, frequency and oxygen concentration.
7. Should have display of at least 6 inches for set parameters and graphical display for measured parameters
8. Should have automatic self test and leak test.
9. Anesthesia machine should be with 3 gas supply system (O<sub>2</sub>, N<sub>2</sub>O, Air) with pipeline connections and reserve cylinder yokes.
10. Gas cylinder (pin indexed) yokes with sturdy clamping bars for easy handling.
11. Pin index yokes for connecting cylinders each for O<sub>2</sub>, N<sub>2</sub>O and air through pipeline.
12. Regulator two each for O<sub>2</sub> and N<sub>2</sub>O. N<sub>2</sub>O should be activated only with oxygen on flow.
13. Should have pressure gauge for all gas inlets including central lines mounted on the front panel for easy visibility
14. Should have audible alarm for O<sub>2</sub> failure
15. N<sub>2</sub>O supply should cut off if O<sub>2</sub> supply fails. (Anti-hypoxic guard).
16. Oxygen and Nitrous oxide should be linked either mechanically or pneumatically to ensure a minimum of 25% oxygen delivery at all times to avoid delivery of hypoxic mixture.
17. Should have dual cascade type flow meter for O<sub>2</sub> and N<sub>2</sub>O and air calibrated in multiple scale.
18. The anesthesia machine should have a master control ON/OFF switch.
19. Provision to mount any two selectable vaporizer with interlocking facility to allow use of only one vaporizer at a time.
20. Iso-flurane vaporizer of newer generation having specifications equivalent to tech 7/8 type to be provided.





21. Non-return cum pressure relief valve when pressure exceeds 120cm of H<sub>2</sub>O.
22. Should have only one common gas outlet.
23. Should provide with oxygen flush switch.
24. Circle absorber with corrugated reusable breathing circuit for closed circuit system with each unit. It should be autoclavable. It should be with ventilator selector switch and circle on/off switch.
25. Should have low flow anesthesia technique.
26. Should have a facility to connect the scavenging system.
27. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.
28. Should have a provision for mounting monitors on top of the machine and with drawers.
29. Patient Monitor SpO<sub>2</sub>, NIBP, ECG, ETCO<sub>2</sub>
30. Should have fiber wheels and Foot brakes.
31. Standard brains circuit
32. Reservoir bag (2liters)
33. Connectors for brains circuit
34. Should be supplied with driver gas hoses with necessary attachments (color coded).
35. Should work in 220-240Vac 50 Hz input power supply.

*Handwritten signature*

## 2. Cardiac Monitor

Technical Specification-Multiparameter Patient monitor	
I.	<b>Multiparameter Monitor</b>
1	Should have TFT/LCD display with at least 10.4 inches with 4 wave forms and numeric display simultaneously. The waveforms should be user selectable.
2	Should be compact & portable with carrying handle
3	Monitor should have in built Lithium-ion type battery for 4 Hrs continuous operation in case of mains failure.
4	Should have keys for quick access to main functions.
5	Should be able to monitor ECG, SpO <sub>2</sub> , NIBP, Respiration Rate for adult, pediatric.
6	3 or 5 Lead ECG monitoring with lethal arrhythmia recognition capability and ST analysis
7	Respiration & Apnea alarm
8	Manual, Auto and STAT mode for NIBP monitoring and ranges should be 20 to 230 mmHg.
9	Pulse Oximeter (SpO <sub>2</sub> ) with Plethysmograph & Pulse strength indicator with variable pitch with change in SpO <sub>2</sub>
10	Side-stream Capnography with display of CO <sub>2</sub> wave form & digital values (ETCO <sub>2</sub> , FiCO <sub>2</sub> , RR).
11	Should have separate volume control for beep sound for QRS and alarm sound.
12	The display setting should have at least 4 user defined setups variable as per applications for flexible use of the monitor in various clinical environments.
13	Monitor should have networking options
14	Should have separate volume control for beep sound for QRS and alarm sound.
15	<b><u>Should provide following accessories</u></b>
	• Reusable adult 3 or 5 lead ECG cable set – 2 nos.
	• Reusable adult and pediatric SpO <sub>2</sub> finger probes – 1 each
	• NIBP cuffs for standard Adult, Obese Adult, Child



16	Facility for last 24 hours trend review facility & facility for patient data entry
17	Equipment performance should not be affected by electromagnetic radiated or conducted through power lines from another device.
18	Should work on 200-240V AC/50Hz with inbuilt rechargeable battery.
19	Should have safety certificate from a competent authority CE/FDA (US)/ STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/test report shall be produced along with the technical bid.
20	Patient monitor trolley

### 3. ERCP Patient Table (OT Table)

#### **Technical Specification of ERCP Patient Table.**

ERCP (Endoscopic Retrograde Cholangiopancreatography) patient table requires a metal-free imaging surface,

motorized adjustments for patient height and position, and robust weight capacity to accommodate a patient and imaging equipment.

Key specifications include a minimum 250kg (550lb) patient load,

height adjustability (e.g., from 570-940mm),

longitudinal and

lateral table movement,

Trendelenburg/reverse

Trendelenburg tilt, and smooth-rolling

lockable casters for mobility and positioning near the C-arm.



**Annexure – 1****PARTICULARS OF THE BIDDER**

<b>Sr. No</b>	<b>Description</b>	<b>Details (to be filled by the responder to the Bid)</b>
1	Name of the company	
2	Official address	
3	Phone No. And Fax No.	
4	Corporate Headquarters Address	
5	Phone No. And Fax No.	
6	Web Site Address	
7	Details of Company's Registration (Please enclose copy of the company registration document)	
8	Name of Registration Authority	
9	Registration Number and Year of Registration	
10	ISO certifications and its validity	
11	GST registration No.	
12	Permanent Account Number (PAN)	
13	Company's Revenue for last 3 years (Year wise)	
14	Company's net worth for the last year	
15	Bank Details (Name, Account no., Branch, IFSC, MICR)	

## Annexure - 2

The Bidder has to quote the rate in the Item Data available online with this bid. Details to be filled up for price bid are as below:

**The price shall be inclusive of all taxes (inclusive of GST) under the relevant Laws of India.**

SL. No	Particular	Amount In Rs. (Inclusive of All the taxes)
1	Total Cost for the Procurement Equipment's for the department of Gastroenterology at Adichunchanagiri Hospital and Research Center, BG Nagara - 571448, Mandya (District)	
<b>Total in Rs and in words -</b>		

Cost related to Supply and Installation as per Items mentioned in the Compliance sheet for technical proposal.

SL. No	Particulars	Total Quantity In nos.	Unit Rate in Rs	Total Cost in Rs.
1	Anesthesia Work Station	1		
2	Cardiac Monitor	1		
3	ERCP Patient Table (OT Table)	1		
<b>Total in Rs</b>				
<b>GST @18% in Rs.</b>				
<b>Grand Total Amount in Rs.</b>				