

SAST/CLPC/ACU-BGSMCH/ 195 /2025-26

Date: 08-12-2025

TENDER NOTIFICATION

The Head, CLPC, Sri Adichunchanagiri Shikshana Trust invites closed tenders from eligible tenderers or bonafide licensed manufacturer (OEM) or their authorized local supplier/ dealer/ distributor in the state of Karnataka for the **Procurement of Instruments for the department of Orthopaedics at BGS MCH Hospital, Nagarur, Nelamangala.** as per section I & II.

01	Name of the work	Procurement of Instruments for the department of Orthopaedics at BGS MCH Hospital, Nagarur, Nelamangala.
02	Last Date for Tender Submission	On or before 31.12.2025 before 5.30 PM

Section-1**Instructions to Tenderers**

- The Tenderer shall submit the bids (Technical & Financial bids) through the mail id: **clpchead@bgscet.ac.in** on or before the last date of tender submission (for any or all list of items) on professional business letterheads only. The details to be printed on the letter head is as follows
 - Tender for Procurement of Instruments for the department of Orthopaedics at BGS MCH Hospital, Nagarur, Nelamangala.
 - Tender Reference number.....[Insert Number]
 - Address to "The HEAD, CLPC, Sri Adichunchanagiri Shikshana Trust, BGSCET Campus, Mahalakshmiapuram, Bengaluru - 560086"
 - The tenderer shall submit the original documents to this office on the last day of submission for verification who prefers to submit the tender through Post can dispatch the same through Registered post / Speed post or Couriers as to reach the above address on or before the due date and time specified in the Tender Notice. Tenders received after the due date and time, for what so ever reasons will not be considered and the authority, Head of CLPC will not be liable or responsible for the same.
- Tender Currency:** Prices shall be quoted in Indian Rupees only.
- AMC/CMC (IF ANY)** is subject to the Sri Adichunchanagiri shikshana trust's norms.
- Warranty:** 3 years.
- Amendment of tender documents:** At any time prior to the deadline of submission of tenders the trust may, for no reason, whether as its own initiative or otherwise modify the tender documents by amendment. Sri Adichunchanagiri Shikshana Trust reserves all the rights to accept, reject, incorporate changes and re-tender without giving any reasons.



- 6) **Documents Comprising the Tender:** Shall attach Brochure, Certification of the product, Bank/account details, PAN, GSTIN, Good Standing Certificate and 02 years of ITR declaration inside the envelope and the company contact details with email id on the in the below mention format in annexure - 1.
- 7) **Tender Prices:** Prices indicated on the price schedule shall be entered separately I.e. the price of the goods, quoted (ex-works, ex-factory, ex-showroom, ex-warehouse, or off-the-shelf, as applicable), including all duties and sales and the other taxes already paid or payable. Any Indian duties, sales and other taxes which will be payable on the goods if the contract is awarded. Conditional tenders will not be considered. The bidder has to give the quotation in the below enclosed format in annexure - 2.
- 8) **Validity of the Bid:** 90 days from the last date of submission of bid.
- 9) **Corrupt or Fraudulent practices:** Sri Adichunchanagiri Shikshana Trust requires that the tenderers, observe the highest standard of ethics during the procurement and execution of such contracts. In purchase of this policy:
 - a) Will reject a proposal for award if it determines the tenderer recommended for award has engaged in corrupt or fraudulent practices in competing for the contract in question;
 - b) Will declare a firm ineligible, either indefinitely or for the stated period of time, to be awarded a university contract if it any time determines that the firm has engaged in corrupt or fraudulent practices in competing for, or in executing, a trust contract.
- 10) **Process to be confidential:** Information relating to the examination, clarification, evaluation, and comparison of tenders and recommendations for the award of contract will not be disclosed to tenderers or any other persons not officially concerned with such process until the award to the successful tenderer has been announced. Any effort by a tenderer to influence the employer's processing of tenders or award decisions may result in rejection of his tender.
- 11) **Clarification of Tenders:** To assist in the examination, evaluation, and comparison of tenders the employer may, at his discretion, ask and tenderer for clarification of his tender, including breakdowns of unit rates. The request for clarification and the response shall be writing or by cable, but no change in the price or substance of the tender shall be sought, offered, or permitted except as required to confirm the correction of arithmetic errors discovered by the employers in the evaluation of the tenders.
- 12) **Delivery:** The successful BIDDER should commence the service as per the tender document/work or purchase order. For any queries or assistance, please write to clpchead@bgscet.ac.in or telephone to +91-8123707324.
- 13) **Penalty Clause:** Non-execution of supply order - for the reasons of failure to supply partially or completely within the stipulated time or any event of breach of contract. In case at any following stages
 - a) For the delayed supply (3 days of grace period) - 5% deduction
 - b) Quantity issues - 5 % deduction
 - c) Quality issues - 10% deduction



Arthroscopy Surgical instruments:

Sl.No.	Name of the item	Specification	Qty	Rate in Rs.	Total Amount in Rs.
1	Hand instruments	All hand instruments should be USFDA/CE/BIS approved should be supplied from the same manufacturer. I should have one piece outer shaft with excellent control over the cutting process.			
	Punches	<ul style="list-style-type: none">• Straight cutting width 3.4 mm, shaft diameter 3 mm, working length 12 cm.• Upturned 15 degrees cutting width 3.4 mm, shaft diameter 3 mm, working length 12 cm.• Jaws curved 90 deg left cutting width 3.4 mm, shaft diameter 3 mm, working length 12 cm and 19 cm.			
	Grasper	<ul style="list-style-type: none">• Aggressive foreign body grasper shaft diameter 3mm, working length 12 cm.			
	Arthroscope probe	Arthroscopy probe should be graduated at 2mm intervals and approximately 18 cm long.			
	Changing rod	Changing rod/ switching stick for knee and shoulder arthroscopy Knee Shoulder			
3	Arthroscopic PCL rasp	The curved tunnel/notchplasty rasp for completing the notchplasty and chamfering of the tibial and femoral tunnel rim.			
4	Arthroscopic Suture Retriever	Diameter 2.75 mm Straight and curved (45 degree left and right) Straight and curved shaft for knee.			
5	Arthroscopic standard Basket forceps	Straight and 15-degree curved up biter left and right Shaft diameter 3.4 mm Bite size standard 1.7 mm.			



6	Arthroscopic bird beak forceps	Straight shaft with 45 degree upcurved jaw and straight jaw with 45 degrees left and right curved handle- autoclavable and reusable.			
7	Arthroscopic fast suture passer	3.2mm (Knee) fast pass suture passer instrument- autoclavable and reusable			
8	Flip cutter for All inside ligament reconstruction	Sizes 7mm, 8mm, 9mm			
9	Arthroscopic Micro-fracture Instrument	Multiple Microfracture Awl angled 30°, 45° and 90° angled. 6mm drill depth Metal Size 3-5 mm			
10	Shoulder Arthroscopy set	<p>Advanced Shoulder Arthroscopy Set including General Instruments, Suture Passing and retrieval devices.</p> <ul style="list-style-type: none">• Hook Probe, 5.0-5.4 mm/Compatible with the set.• Tissue penetrator & Suture Passer, 45 Degrees and 22- 30 degrees up tip, 2.75mm/ Compatible with the set.• Suture Retriever, 3.4 mm/ Compatible with the set, straight.• Shoulder debridement Rasp, 20–30-degree angle / compatible with the set.• Glenoid Rasp, 20 Degree tip /Compatible with the set.• Bankart Rasp, 90 Degree tip / Compatible with the set.• Knot pusher closed end.• Shoulder Tissue Elevator, 15°/ Compatible with the set.• Shoulder Tissue Elevator, 20-			

Section-2**Technical Specification****Arthroscopy Set:**

Sl.No.	Name of the item	Specification	Qty	Rate in Rs.	Total Amount in Rs.
1	Arthroscope	Autoclavable Wide angle forward oblique and lateral scope Angle of view= 30/70 degrees (To be decided by User) Diameter 4mm (Diameter 2.7mm for Ankle) Length 14 cm or more Fiber optic light transmission incorporated Standard ocular window for coupling camera head			
2	Arthroscope sheath rotatable (compatible with Arthroscope)	Autoclavable Diameter and working length compatible with arthroscope. With 2 rotatable stopcocks/ snap-in coupling mechanism.			
3.	Integrated Camera control System	4K Integrated Camera System with LED/OLED Light Source The system should be Full HDTV Endoscope Video Camera Optical zoom lens with desirable range (mm) F-12-35. 3 Camera Head. HD camera head should have resolution of minimum 3840*2160. Camera head should be having Housing and hermetically Sealed for autoclaving. Camera head should have programmable buttons and inbuilt zoom facility. White balance should be automatically triggered by controller front and camera head.			



		Light Guide fiber optic cables (diameter- 5 mm or more and length at least 10 ft)- should be extremely heat resistant. Fully Integrated Advance Image Management System Should have USB port for recording.			
4	High resolution monitor	32 inch - 4K Medical grade Monitor with resolution of 3840*2160.			
5	Shaver System (Shaver console)	The console screen should capture all information pertaining to minimum and maximum, should provide control for push switches for increasing and decreasing speed setting. There are three modes: standard, efficient and aggressive. Should provide variable RPM (Range:100 to 8000) as per the blade or burr. Motor should offer forward, reverse oscillation mode for resection. Maximum RPM forward/reverse 8000 RPM or more Oscillation 3000 or more			
6	Shaver Hand Piece and Shaver foot pedal	Autoclavable shaver hand piece with hand control Hand piece should have suction control lever and should be detachable. Two different hand piece should be provided for large and short hand procedures.			



10	Equipment Cart with Articulating Arm for Arthro-carts.	<ul style="list-style-type: none">• Video cart should be of same make and should have shock proof powder coating material used.• Should have detachable cable guards.• Should have isolating transformer 2000 VA with earth leakage• Allow mounting position for monitor mount• Should have cable winding aid.• Shaver Console & Footswitch.• Should also have foot pedal holder, camera holder and fluid bag holder.• 1 Fluid Management System.• Size of drawers should be compatible with the size of the equipment• Lockable wheels			
11	Semi sharp obturator	<ul style="list-style-type: none">• Use with arthroscopy sheath, working length compatible with the set.			
Total amount in Rs.					
GST%					
Grand Total Amount					



7	RF Ablation System	<ul style="list-style-type: none">• Automatic detection of connected probe• Three button controlled with ablation, coagulation, ablation power setting, metal proximity detection-suspends power when close to metal objects.• Protection against electric shock according to IEC60601-1.			
8	Arthroscopy fluid management pump/Irrigation System	<ul style="list-style-type: none">• Operation mode, inflow/outflow, flowrate more than 1500ml/min.• Single button setting for all arthroscopic procedures• Pressure setting 10-120 (Increments of 5 value)• Shaver and RF interface function should be available.• Automatic instrument recognition system• In addition to hand control features, the foot switch can also be used to control drain, clear and vacuum functions.			
9	Recording System	<ul style="list-style-type: none">• Full HD• Should have LED monitor for live view, play back or GUI.• Built in hard drive (500GB or more) should be included for recording data back up and user should also can record to USB memory and USB hard drives.			

Handwritten signature

		30deg/ compatible with the set. • Rotator cuff Grasper. • Foreign body grasper- aggressive, straight jaws, sheath diameter 3.5 mm, straight, working length 19 cm. • Suture grasper- straight jaws sheath diameter 3 mm straight, working length 14 cm.			
Total amount in Rs.					
GST%					
Grand Total Amount					

Hip Surgery Instruments

SL. NO	NAME	Qty	Rate in Rs.	Total Amount in Rs.
1	Anterior column Hohmann	4		
2	Cobb elevator	2		
3	Charnley's retractor	2		
4	Nibblers large & Medium size	2		
5	Long toothed forceps	4		
6	Murphy's skid	2		
7	Self-retaining large retractors	4		
8	Penny walker long periosteum	2		

	(wooden handle)			
Total amount in Rs.				
GST@				
Grand Total Amount (Inclusive GST)				

Knee Surgery Instruments

SL. NO	NAME	Qty	Rate in Rs.	Total Amount in Rs.
1	Posterior 90° retractors- PCL retractors	2		
2	Curved osteotomes	10mm - 2 15mm - 2		
3	Medium sized Hohmann spikes	4		
4	Medium size nibbler	2		
5	Medium size wooden handle periosteum	2		
6	Medium size curved gouge	2		
Total Amount in Rs.				
GST@ in Rs.				
Grand Total Amount in Rs.				



Shoulder Instruments

SL. NO	NAME	Qty	Rate in Rs.	Total Amount in Rs.
1	Fukuda retractors	2		
2	Hohmann's spike	2		
Total Amount in Rs.				
GST@ in Rs.				
Grand Total Amount in Rs.				

Small bone fracture Instruments:

SL. NO	NAME	Qty	Rate in Rs.	Total Amount in Rs.
1	Spikes – Small size	6		
2	Periosteum elevator	2		
3	Small size levers	4		
4	Hintermann's K-wise distractor	2		
Total Amount in Rs.				
GST@ in Rs.				
Grand Total Amount in Rs.				





General Instruments

SL. NO	NAME	Qty	Rate in Rs.	Total Amount in Rs.
1	Retractors – Small, Medium & Large	4 Set		
2	Standard Osteotomes - Small, Medium & Large	4		
3	90° Artery forceps	4		
4	Straight Kocher's	6		
5	Straight artery	6		
6	SS – Wire passing instruments	2		
7	Power drill- dual trigger , 270-1200 rpm, forward & reverse mode, battery operated, battery charger, cordless, with drilling, reaming, sagittal attachmets	1		
Total Amount in Rs.				
GST@ in Rs.				
Grand Total Amount in Rs.				

Spine Instruments

SL. NO	NAME	Qty	Rate in Rs.	Total Amount in Rs.
1	GELPI retractor long	2		
2	Bayonet dura retractor	1		

Handwritten signature



3	Ball point hook	1		
4	Ramani hinges self-retaining micro discectomy retractor (6 blade & 3 Hook)	1		
5	Cobbs periosteal elevator with big blade	1		
6	Codman bone nibbler (10mm)	1		
7	spinous process cutter Big	1		
8	Bolster (Silicon)	1		
9	Silicon Pads for Elbow & Knee	4		
10	Gardner Wells tongs Cervical Traction	2		
11	Ozone machine	1		
Total Amount in Rs.				
GST@ in Rs.				
Grand Total Amount in Rs.				

ENDOSCOPIC SPINE INSTRUMENT

SL. NO	INSTRUMENTS	Qty	Rate in Rs.	Total Amount in Rs.
1	Steno Scope Length 177MM	1		
2	Sealing Top Aadaptor	1		
3	Sealing Cap	1		
4	Dilator	1		



5	Working Sleeve	1
6	Flushing Adapter work	1
7	Kerrison punch 5.5* 4.5MM WL 380MM	2
8	Punch 4MM WL 290MM	2
9	Rongeur 4MM WL 290MM	2
10	Annulotome 2.5 MM WL 350MM	2
11	Dissector	1
12	Palpation Hook	1
13	Face miller	1
14	Aluminum Tray	1
15	Head Light Set	1

Total Amount in Rs.

GST@ in Rs.

Grand Total Amount in Rs.

Annexure – 1**PARTICULARS OF THE BIDDER**

Sr. No	Description	Details (to be filled by the responder to the Bid)
1	Name of the company	
2	Official address	
3	Phone No. And Fax No.	
4	Corporate Headquarters Address	
5	Phone No. And Fax No.	
6	Web Site Address	
7	Details of Company's Registration (Please enclose copy of the company registration document)	
8	Name of Registration Authority	
9	Registration Number and Year of Registration	
10	ISO certifications and its validity	
11	GST registration No.	
12	Permanent Account Number (PAN)	
13	Company's Revenue for last 3 years (Year wise)	
14	Company's net worth for the last year	
15	Bank Details (Name, Account no., Branch, IFSC, MICR)	




Annexure – 2

The Bidder has to quote the rate in the Item Data available online with this bid. Details to be filled up for price bid are as below:

The price shall be inclusive of all taxes (inclusive of GST) under the relevant Laws of India.

SL. No	Particular	Amount In Rs. (Inclusive of All the taxes)
1	Total Cost for the Procurement of Instruments for the department of Orthopaedics at BGS MCH Hospital, Nagarur, Nelamangala.	
Total in Rs and in words –		


THE HEAD
CLPC, ACU
 Mahalakshmiapuram,
 Bengaluru - 560 086.