



### **TENDER NOTIFICATION**

The Head, CLPC, Sri Adichunchanagiri Shikshana Trust invites closed tenders from eligible tenderers or bonafide licensed manufacturer (OEM) or their authorized local supplier/ dealer/ distributor in the state of Karnataka for the **Procurement of baby warmers, t-piece, and Video laryngoscope for the NICU department of Pediatrics Adichunchanagiri Hospital and Research Center, BG Nagara - 571448, Mandyā (District)** as per section I & II.

01	Name of the work	Procurement of baby warmers, t-piece, and Video laryngoscope for the NICU department of Pediatrics Adichunchanagiri Hospital and Research Center, BG Nagara - 571448, Mandyā (District)
02	Last Date for Tender Submission	On or before <b>06.02.2026</b> 5.30 PM (Tender Submission to this mail id <a href="mailto:clpcheadtender@bgscet.ac.in">clpcheadtender@bgscet.ac.in</a> is Mandatory).

#### **Section-1**

#### **Instructions to Tenderers**

- 1) The Tenderer shall submit the bids (Technical & Financial bids) through the mail id: [clpcheadtender@bgscet.ac.in](mailto:clpcheadtender@bgscet.ac.in) on or before the last date of tender submission (for any or all list of items) on professional business letterheads only. The details to be printed on the letter head is as follows
  - i) Tender for the Procurement of baby warmers, t-piece, and Video laryngoscope for the NICU department of Pediatrics Adichunchanagiri Hospital and Research Center, BG Nagara - 571448, Mandyā (District)
  - ii) Tender Reference number.....[Insert Number]
  - iii) Address to "The HEAD, CLPC, Sri Adichunchanagiri Shikshana Trust, BGSCET Campus, Mahalakshmiapuram, Bengaluru – 560086"
  - iv) The tenderer shall submit the original documents to this office on the last day of submission for verification who prefers to submit the tender through Post can dispatch the same through Registered post / Speed post or Couriers as to reach the above address on or before the due date and time specified in the Tender Notice. Tenders received after the due date and time, for what so ever reasons will not be considered and the authority, Head of CLPC will not be liable or responsible for the same.
- 2) **Tender Currency:** Prices shall be quoted in Indian Rupees only.
- 3) **AMC/CMC (IF ANY)** is subject to the Sri Adichunchanagiri shikshana trust's norms.
- 4) **Warranty:** 3 Years.
- 5) **Amendment of tender documents:** At any time prior to the deadline of submission of tenders the trust may, for no reason, whether as its own initiative or otherwise modify the tender documents by amendment. Sri Adichunchanagiri Shikshana Trust reserves all the rights to accept, reject, incorporate changes and re-tender without giving any reasons.

1 | Page



- 6) **Documents Comprising the Tender:** Shall attach Brochure, Certification of the product, Bank/account details, PAN, GSTIN, Good Standing Certificate and 02 years of ITR declaration inside the envelope and the company contact details with email id on the in the below mention format in annexure - 1.
- 7) **Tender Prices:** Prices indicated on the price schedule shall be entered separately i.e. the price of the goods, quoted (ex-works, ex-factory, ex-showroom, ex-warehouse, or off-the-shelf, as applicable), including all duties and sales and the other taxes already paid or payable. Any Indian duties, sales and other taxes which will be payable on the goods if the contract is awarded. Conditional tenders will not be considered. The bidder has to give the quotation in the below enclosed format in annexure - 2.
- 8) **Validity of the Bid:** 90 days from the last date of submission of bid.
- 9) **Corrupt or Fraudulent practices:** Sri Adichunchanagiri Shikshana Trust requires that the tenderers, observe the highest standard of ethics during the procurement and execution of such contracts. In purchase of this policy:
  - a) Will reject a proposal for award if it determines the tenderer recommended for award has engaged in corrupt or fraudulent practices in competing for the contract in question;
  - b) Will declare a firm ineligible, either indefinitely or for the stated period of time, to be awarded a university contract if it any time determines that the firm has engaged in corrupt or fraudulent practices in competing for, or in executing, a trust contract.
- 10) **Process to be confidential:** Information relating to the examination, clarification, evaluation, and comparison of tenders and recommendations for the award of contract will not be disclosed to tenderers or any other persons not officially concerned with such process until the award to the successful tenderer has been announced. Any effort by a tenderer to influence the employer's processing of tenders or award decisions may result in rejection of his tender.
- 11) **Clarification of Tenders:** To assist in the examination, evaluation, and comparison of tenders the employer may, at his discretion, ask and tenderer for clarification of his tender, including breakdowns of unit rates. The request for clarification and the response shall be writing or by cable, but no change in the price or substance of the tender shall be sought, offered, or permitted except as required to confirm the correction of arithmetic errors discovered by the employers in the evaluation of the tenders.
- 12) **Delivery:** The successful BIDDER should commence the service as per the tender document/work or purchase order. For any queries or assistance, please write to [clpchead@bgscet.ac.in](mailto:clpchead@bgscet.ac.in) or telephone to +91- 8123707324.
- 13) **Penalty Clause:** Non-execution of supply order – for the reasons of failure to supply partially or completely within the stipulated time or any event of breach of contract. In case at any following stages
  - a) For the delayed supply (3 days of grace period) – 5% deduction
  - b) Quantity issues – 5 % deduction
  - c) Quality issues – 10% deduction

**Section 2****Technical Specification**

SL. No	Description	QTY
1	Baby Warmers	03
2	T piece resuscitator	01
3	Video laryngoscope	01
4	KMC Chair	04

**Technical Specification****Baby warmer**

1. The unit should be made of mild steel tubular structure pre-treated and powder coated
2. Heater Rotation  $\pm 90^\circ$  to the side to facilitate X-ray procedures
3. Bed Tilt should be  $\pm 15^\circ$  Trendelenburg and Reverse Trendelenburg, continuous tilt
4. Mattress density should be approx. 21-25 kg/m<sup>3</sup> and removable, washable, water proof cover
5. Mattress density should be approx. 21-25 kg/m<sup>3</sup> and removable, washable, water proof cover
6. System Control specifications
7. Should have microprocessor-based heater control and manual modes of operation
8. Should have user friendly control panel with large easy to read LED displays for actual (patient and air temperature) and set temperatures.
9. Should have Quartz Infrared Heater with parabolic reflector for uniform heat Radiation.
10. The heater unit should be protected by a suitable grill.
11. The probes should be detachable type.
12. Should have calibration free temperature sensors.
13. Should have alarms with visual indicators for the following
14. Should have alarms with visual indicators for the following
  - Temp high
  - Temp low
  - Probe failure
  - Power Heater failure
15. The heater should automatically cut off at 38 degree Celsius irrespective of the set parameters.
16. Should have an examination light with ON/OFF switch.
17. Should work with input 200 to 240Vac 50 Hz supply
18. Should have 0-650 W heater output
19. Heater output should be adjustable from 0 - 100% in 5% increments



20. servo Control should be between 30 - 38°C in increments of 0.1°C
21. Manual Mode should Indicate manual mode heat selection range from: 0-100% in 5% increments
22. Temperature Measurement Accuracy specification:  $\pm 0.3^\circ\text{C}$  @ 30°C to 40°C
23. Temperature Display Resolution specification:  $\pm 0.1^\circ\text{C}$
24. Temperature Probe Accuracy specification:  $\pm 0.1^\circ\text{C}$  @ 30°C to 42°C
25. Humidity range: 30 to 95% RH
26. Should have safety certificate from a competent authority CE.

## Technical Specification

### Infant T- Piece Resuscitator: -

1. The device should be ideal for use in Labor room, NICU and during transportation.
2. It should be powered by gas flow with no electrical or battery operation.
3. It should be light weight and easy to handle.
4. In-built Pressure Gauge ( manometer ) to set & indicate delivery of PIP & PEEP
  - a. Manometer range -10 to 70 cmH20 (mbar)
  - b. Manometer accuracy:  $\pm 2.0\%$  full scale deflection.
5. Peak Inspiratory Pressure ( PIP ) at 8 lpm : 3 to 59 cmH20
6. Positive End Expiratory Pressure ( PEEP ) at 8lpm : 0 cmH2O to 8 cmH2O
7. Safety provision with adjustable Pressure Relief Valve for maximum limiting
8. Maximum pressure relief at 8LPM : 5 to 70cm H2O
9. Delivered oxygen up to 100% depending on gas supply.
10. Unit should be compatible with neonatal mask & endotracheal tube
11. The patient T -Piece should have port for surfactant delivery
12. Should be compatible for use with heated humidifier.
13. Spiral Heated wire circuit with integrated T-Piece compatible for use with heated humidifier should be supplied with system.



14. Operating and storage limits -10°C to 50°C and Up to 90% relative humidity

15. Certification: Manufacturer should have US FDA Approval.

16. It should meet safety standards IEC 60601-1, EN 60601-1. ISO standards for medical devices: ISO 13485 & ISO 10651-5

17. Supply line – Two sets with each

18. Test lung- One

19. T Piece Resuscitation Circuits with PEEP valve – 1 piece

20. Resuscitation Masks of Premature size (medium) - 1 Piece


21. Resuscitation Masks of Micro Premature size (small) - 1 Piece

## **TECHNICAL SPECIFICATIONS**

### **Video Laryngoscope: -**

1. Should be a video laryngoscope convenient for tracheal intubation.
2. Should have a camera for live Image capturing.
3. Should have LED light illumination
4. Should have color Image display facility LCD/TFT display
5. Should have provision to insert all sizes of endotracheal tube
6. Should have a provision to introduce all sizes of suction catheters
7. Should have water proof protection
8. Should be supplied with rechargeable battery and provision for re-charge.
9. Should have a battery backup facility of minimum30 min
10. Should have all blade sizes/adjustable for adult and paediatric laryngoscopy. If the blades are disposable, should supply 50nos. of blades compatible for both adult and paediatric along with each unit.
11. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.

5 | Page

**Technical Specifications for Kanagroo Mother Care Chair**

<b>Material</b>	Mild Steel
<b>Type</b>	Kanagroo Mother Care Chair
<b>Usage</b>	Hospital
<b>Color</b>	Blue
<b>Specific Usage</b>	Mother Care
<b>Finishing</b>	Paint Coated
<b>Furniture Type</b>	Hospital Chair
<b>Metals Type</b>	Steel
<b>Size</b>	7x2 feet
<b>Feature</b>	Adjustable Height



6 | Page



## Annexure - 1

## PARTICULARS OF THE BIDDER

Sr. No	Description	Details (to be filled by the responder to the Bid)
1	Name of the company	
2	Official address	
3	Phone No. And Fax No.	
4	Corporate Headquarters Address	
5	Phone No. And Fax No.	
6	Web Site Address	
7	Details of Company's Registration (Please enclose copy of the company registration document)	
8	Name of Registration Authority	
9	Registration Number and Year of Registration	
10	ISO certifications and its validity	
11	GST registration No.	
12	Permanent Account Number (PAN)	
13	Company's Revenue for last 3 years (Year wise)	
14	Company's net worth for the last year	
15	Bank Details (Name, Account no., Branch, IFSC, MICR)	

7 | Page

**Annexure - 2**

The Bidder has to quote the rate in the Item Data available online with this bid. Details to be filled up for price bid are as below:

**The price shall be inclusive of all taxes (inclusive of GST) under the relevant Laws of India.**

<b>SL. No</b>	<b>Particular</b>	<b>Amount In Rs. (Inclusive of All the taxes)</b>
1	Total cost of the Procurement of baby warmers, t-piece, and Video laryngoscope for the NICU department of Pediatrics Adichunchanagiri Hospital and Research Center, BG Nagara - 571448, Mandya (District)	
<b>Total in Rs and in words –</b>		

**Cost related to Supply and Installation as per Items mentioned in the Compliance sheet for technical proposal.**

<b>SL. No</b>	<b>Particulars</b>	<b>Total Quantity In nos.</b>	<b>Unit Rate in Rs</b>	<b>Total Cost in Rs.</b>
1	Baby Warmers	03		
2	T piece resuscitator	01		
3	Video laryngoscope	01		
4	KMC Chair	04		
<b>Total in Rs</b>				
<b>GST@ % in Rs.</b>				
<b>Grand Total Amount in Rs.</b>				