

SAST/CLPC/ACU-AHRC/337/2025-26

Date: 16-03-2026

**TENDER NOTIFICATION**

The Head, CLPC, Sri Adichunchanagiri Shikshana Trust invites closed tenders from eligible tenderers or bonafide licensed manufacturers (OEM) or their authorized local supplier/ dealer/ distributor in the state of Karnataka for the **Procurement of Equipment's for the Department of Dermatology at AHRC, BG Nagara-571448.** as per section I & II.

01	Name of the work	Procurement of Equipment's for the Department of Dermatology at AHRC, BG Nagara-571448
02	Last Date for Tender Submission	On or before 25.03.2026 before 5.30 PM (Tender Submission to <a href="mailto:clpcheadtender@bgscet.ac.in">clpcheadtender@bgscet.ac.in</a> is Mandatory)

**Section-1****Instructions to Tenderers**

- 1) The Tenderer shall submit the bids (Technical & Financial bids) through the mail id: **clpcheadtender@bgscet.ac.in** on or before the last date of tender submission (for any or all list of items) on professional business letterheads only. The details to be printed on the letter head is as follows
  - i) Tender for Procurement of Equipment's for the Department of Dermatology at AHRC, BG Nagara-571448
  - ii) Tender Reference number.....[Insert Number]
  - iii) Address to "The HEAD, CLPC, Sri Adichunchanagiri Shikshana Trust, BGSCET Campus, Mahalakshmiapuram, Bengaluru - 560086"
  - iv) The tenderer shall submit the original documents to this office on the last day of submission for verification who prefers to submit the tender through Post can dispatch the same through Registered post / Speed post or Couriers as to reach the above address on or before the due date and time specified in the Tender Notice. Tenders received after the due date and time, for what so ever reasons will not be considered and the authority, Head of CLPC will not be liable or responsible for the same.
- 2) **Tender Currency:** Prices shall be quoted in Indian Rupees only.
- 3) **AMC/CMC (IF ANY)** is subject to the Sri Adichunchanagiri shikshana trust's norms.
- 4) **Warranty:** 3 Years.
- 5) **Amendment of tender documents:** At any time prior to the deadline of submission of tenders the trust may, for no reason, whether as its own initiative or otherwise modify the tender documents by amendment. Sri Adichunchanagiri Shikshana Trust reserves

all the rights to accept, reject, incorporate changes and re-tender without giving any reasons.

- 6) **Documents Comprising the Tender:** Shall attach Brochure, Certification of the product, Bank/account details, PAN, GSTIN, Good Standing Certificate and 02 years of ITR declaration inside the envelope and the company contact details with email id on the in the below mention format in annexure - 1.
- 7) **Tender Prices:** Prices indicated on the price schedule shall be entered separately I.e. the price of the goods, quoted (ex-works, ex-factory, ex-showroom, ex-warehouse, or off-the-shelf, as applicable), including all duties and sales and the other taxes already paid or payable. Any Indian duties, sales and other taxes which will be payable on the goods if the contract is awarded. Conditional tenders will not be considered. The bidder has to give the quotation in the below enclosed format in annexure - 2.
- 8) **Validity of the Bid:** 90 days from the last date of submission of bid.
- 9) **Corrupt or Fraudulent practices:** Sri Adichunchanagiri Shikshana Trust requires that the tenderers, observe the highest standard of ethics during the procurement and execution of such contracts. In purchase of this policy:
  - a) Will reject a proposal for award if it determines the tenderer recommended for award has engaged in corrupt or fraudulent practices in competing for the contract in question;
  - b) Will declare a firm ineligible, either indefinitely or for the stated period of time, to be awarded a university contract if it any time determines that the firm has engaged in corrupt or fraudulent practices in competing for, or in executing, a trust contract.
- 10) **Process to be confidential:** Information relating to the examination, clarification, evaluation, and comparison of tenders and recommendations for the award of contract will not be disclosed to tenderers or any other persons not officially concerned with such process until the award to the successful tenderer has been announced. Any effort by a tenderer to influence the employer's processing of tenders or award decisions may result in rejection of his tender.
- 11) **Clarification of Tenders:** To assist in the examination, evaluation, and comparison of tenders the employer may, at his discretion, ask and tenderer for clarification of his tender, including breakdowns of unit rates. The request for clarification and the response shall be writing or by cable, but no change in the price or substance of the tender shall be sought, offered, or permitted except as required to confirm the correction of arithmetic errors discovered by the employers in the evaluation of the tenders.
- 12) **Delivery:** The successful BIDDER should commence the service as per the tender document/work or purchase order. For any queries or assistance, please write to [clpthead@bgscet.ac.in](mailto:clpthead@bgscet.ac.in) or telephone to +91- 8123707324.
- 13) **Penalty Clause:** Non-execution of supply order – for the reasons of failure to supply partially or completely within the stipulated time or any event of breach of contract. In case at any following stages
  - a) For the delayed supply (3 days of grace period) – 5% deduction
  - b) Quantity issues – 5 % deduction
  - c) Quality issues – 10% deduction

**Section-2****Technical Specification**

Sl. No	Description	Uom	Qty	Rate Per unit in Rs.	Total amount per unit
1	Excimer Laser	Nos	1		
2	Multility Phototherapy Unit	Nos	1		
3	Microdermabrasion Unit	Nos	1		
4	Liquid Nitrogen Cryo Unit	Nos	1		
5	Derma Chair	Nos	1		
6	Hairlite Equipment	Nos	1		
<b>Total Amount</b>					
<b>INR in Rupees</b>					




### Technical Specifications UV Excimer Laser

- 1 Wavelength: 248 nm
- 2 Pulse energy: 750mJ @248 nm (or higher)
- 3 Pulse Repetition Rate: 10 Hz or more
- 4 Average Power: 7 W or higher
- 5 Pulse Width: ~20 ns (FWHM)
- 6 Pulse – to –Pulse Stability: 1% or better
- 7 Beam orientation: Vertical
- 8 Beam Dimension: ~ 12 x 28 mm<sup>2</sup> with a margin of  $\pm 20\%$
- 9 Beam Divergence: ~3 x 1 mrad<sup>2</sup> or smaller
- 10 Electrical power: 220 V Single Phase Operation
- 11 Weight: Less than 350 kg
- 12 Capable of Remote Operation with PC interface
- 13 Dimensions: less than  $\leq 200 \times 500 \times 1000$  mm<sup>3</sup>
- 14 Air Cooled System
- 15 One KrF premix gas cylinder ~50 litres
- 16 One He gas cylinder ~50 litres
- 17 Energy Range: 0.2 mJ to 1 J
- 18 Wavelength Range: 190 to 266 nm
- 19 Active Area Diameter: 50 mm
- 20 Maximum Average Power: ~35 W (with heat sink)
- 21 Damage Threshold: > 500mJ/cm<sup>2</sup>
- 22 Display Unit: LED/LCD display with USB interface



## Technical specification for multity phototherapy unit

1. The unit has 12 narrow band UVB tubes.
2. Electronic ballast for the tubes
3. The unit is provided with imported mirror type reflectors.
4. Unit has provision to tilt all the panels from the sides with spring movement.
5. Independent ON OFF control for each panel.
6. Cabinet made of high quality steel & powder coated.
7. Four line LCD display for the control panel.
8. Lock & key provided for the control panel.
9. Phototherapy patient data card is also available.

A handwritten signature in black ink, appearing to be 'H. Anur', is written over a horizontal line.

## **Technical Specifications for Micro dermabrasion**

### **Units:**

- Voltage/Frequency: 110V–240V AC, 50-60 Hz.
- Power Consumption: Typically ranges from 45W to 100W, with some heavy-duty units up to 320W.
- Suction Power: 0–0.82 bar (approx. 615 mmHg) or 65–80 kPa.
- Air Flow: Approximately 28 L/min.
- Dimensions: Ranging from compact (approx. 30x20x15 cm) to larger cabinet units.
- Weight: Portable units are often 1–5 kg, while larger, more robust units can be 10–20 kg.
- Material: Usually durable ABS plastic or metal.
- Components: 2-in-1, 3-in-1, or 5-in-1 combinations, often including diamond-tip wands, suction hoses, spray bottles, and replaceable filters.
- Tip Variations: Sets typically include fine (250+ grit), medium (150–200 grit), and coarse (80–120 grit) tips for different skin types.

*Handwritten signature*

**Technical specification Liquid Nitrogen cryo Production Unit**

- Production Capacity: Ranging from small units producing 1-10 liters per day to industrial units producing 60 liters per day or more.
- Purity: Typically 99% nitrogen purity.
- Technology: Generally uses GM (Gifford-McMahon) or Stirling Cycle for cooling.
- Pressure: Adjustable pressure boost, often up to 4 bar, with HMI touch screen control.
- Compressor: Oil-free, air-cooled compressor.
- Capacity: Ranges from 3 liters to over 1200 liters (e.g., 340L, 550L, 1200L models).
- Construction: High-strength aluminum for portable units, stainless steel (SS304/SS316L) for large bulk storage.
- Evaporation Rate: Low evaporation rate, often <1% per day for large vessels, or 0.11 - 0.37 L/day for smaller containers.
- Neck Diameter: Varies from 50mm (small) up to 8.5 inches or more (wide mouth).
- Safety Features: Pressure relief valves, vacuum-insulated, and often paired with oxygen deficiency monitors.



**Technical specification for derma chair.**

1. Length - 1840mm
2. Width - 560mm
3. Hight - 630-850mm
4. Back section- 60degree
5. Leg section - 0-75degree
6. Trendelendurg - 25degree
7. Lateral tilt - 20degree
8. CPR function - optional
9. Battery - optional
10. Foot operated remote- optional
11. Metal - Epoxy powder coating.
12. Load capacity-100-150kg
13. Chair weight -80-90kg.



**Technical specification for hair lite equipment:**

- Electrical: 100-120 VAC, 20A, 50/60 Hz or 220-240 VAC, 10A, 50/60 Hz
- Weight: 63 Kg (139 lbs)
- Dimensions: 35 cm wide x 40 cm deep x 110 cm high (14" x 16" x 43.3")
- SHR Pro (Super Hair Removal) Mode:
  - Technology: In-Motion AFT (Advanced Fluorescence Technology)
  - Wavelength: NIR (Near-Infrared)
  - Pulse Width: 8 msec
  - Spot Size: 5 cm<sup>2</sup>
  - Frequency: 3 Hz
- Technology: IPL (Intense Pulsed Light) with Sapphire light window for ice cooling
- Indications: Hair removal and long-lasting reduction in hair regrowth
- Functions: High mode, Normal mode, Fast mode, Auto-play, Ice cooling
- Components: Main unit, power adaptor/cord, razor,



**Annexure – 1****PARTICULARS OF THE BIDDER**

<b>Sr. No</b>	<b>Description</b>	<b>Details (to be filled by the responder to the Bid)</b>
1	Name of the company	
2	Official address	
3	Phone No. And Fax No.	
4	Corporate Headquarters Address	
5	Phone No. And Fax No.	
6	Web Site Address	
7	Details of Company's Registration (Please enclose copy of the company registration document)	
8	Name of Registration Authority	
9	Registration Number and Year of Registration	
10	ISO certifications and its validity	
11	GST registration No.	
12	Permanent Account Number (PAN)	
13	Company's Revenue for last 3 years (Year wise)	
14	Company's net worth for the last year	
15	Bank Details (Name, Account no., Branch, IFSC, MICR)	



**Annexure – 2**

The Bidder has to quote the rate in the Item Data available online with this bid. Details to be filled up for price bid are as below:

**The price shall be inclusive of all taxes (inclusive of GST) under the relevant Laws of India.**

SL. No	Particular	Amount In Rs. (Inclusive of All the taxes)
1	Procurement of Equipment's for the Department of Dermatology at AHRC, BG Nagara-571448	
<b>Total in Rs and in words -</b>		



**THE HEAD  
CLPC, ACU  
Mahalakshmpuram,  
Bengaluru - 560 080.**